

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062620\
 Data File : VN062166.D
 Acq On : 27 Jun 2020 2:40
 Operator : JC/MD
 Sample : L2817-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-01-062420

Quant Time: Jun 27 06:50:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	214131	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	405896	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	406252	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	152092	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	152293	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
35) Dibromofluoromethane	7.55	113	128382	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
50) Toluene-d8	10.06	98	525126	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	12.38	95	175413	49.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.76%	
Target Compounds						
16) Acetone	3.78	43	10060	6.724	ug/l	98
18) Methyl Acetate	4.28	43	4241	1.775	ug/l #	82
20) Methylene Chloride	4.51	84	9684	3.308	ug/l	95
43) Isopropyl Acetate	8.13	43	49134	8.922	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062620\
Data File : VN062166.D
Acq On : 27 Jun 2020 2:40
Operator : JC/MD
Sample : L2817-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-01-062420

Quant Time: Jun 27 06:50:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

