

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062620\
 Data File : VN062167.D
 Acq On : 27 Jun 2020 3:04
 Operator : JC/MD
 Sample : L2815-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-062420

Quant Time: Jun 27 06:52:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	216117	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	410412	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	399631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	146504	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	152543	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
35) Dibromofluoromethane	7.55	113	128431	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
50) Toluene-d8	10.06	98	528340	50.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.40%	
62) 4-Bromofluorobenzene	12.38	95	172260	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
Target Compounds						
16) Acetone	3.79	43	11885	8.573	ug/l	96
18) Methyl Acetate	4.29	43	4195	1.740	ug/l #	66
20) Methylene Chloride	4.50	84	9406	3.175	ug/l	92
43) Isopropyl Acetate	8.13	43	46698	8.386	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062620\
Data File : VN062167.D
Acq On : 27 Jun 2020 3:04
Operator : JC/MD
Sample : L2815-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-062420

Quant Time: Jun 27 06:52:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

