

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062620\
 Data File : VN062170.D
 Acq On : 27 Jun 2020 4:17
 Operator : JC/MD
 Sample : PB129754ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129754ZHE#01

Quant Time: Jun 27 07:02:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	210784	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	410452	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	406286	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	151106	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	152209	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
35) Dibromofluoromethane	7.55	113	126916	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
50) Toluene-d8	10.06	98	528662	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
62) 4-Bromofluorobenzene	12.38	95	174413	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
Target Compounds						
16) Acetone	3.78	43	16241	13.662	ug/l	99
18) Methyl Acetate	4.29	43	4863	2.068	ug/l #	46
20) Methylene Chloride	4.50	84	18946	6.790	ug/l	98
43) Isopropyl Acetate	8.13	43	58740	10.548	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062620\
Data File : VN062170.D
Acq On : 27 Jun 2020 4:17
Operator : JC/MD
Sample : PB129754ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129754ZHE#01

Quant Time: Jun 27 07:02:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

