

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062620\
 Data File : VN062172.D
 Acq On : 27 Jun 2020 5:05
 Operator : JC/MD
 Sample : PB129754ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129754ZHE#03

Quant Time: Jun 27 07:06:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	195780	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	383265	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	383663	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	134289	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	143251	52.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.28%	
35) Dibromofluoromethane	7.55	113	119314	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
50) Toluene-d8	10.06	98	494349	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
62) 4-Bromofluorobenzene	12.38	95	162727	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
Target Compounds						
16) Acetone	3.78	43	17439	16.437	ug/l	93
18) Methyl Acetate	4.28	43	5564	2.548	ug/l #	77
20) Methylene Chloride	4.51	84	29601	11.570	ug/l	99
43) Isopropyl Acetate	8.13	43	55041	10.584	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062620\
Data File : VN062172.D
Acq On : 27 Jun 2020 5:05
Operator : JC/MD
Sample : PB129754ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129754ZHE#03

Quant Time: Jun 27 07:06:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

