

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062620\
 Data File : VN062173.D
 Acq On : 27 Jun 2020 5:29
 Operator : JC/MD
 Sample : PB129754ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129754ZHE#04

Quant Time: Jun 27 07:09:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	202680	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	394918	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	395213	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	145693	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	146882	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
35) Dibromofluoromethane	7.55	113	123301	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
50) Toluene-d8	10.06	98	513351	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
62) 4-Bromofluorobenzene	12.38	95	172124	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
Target Compounds						
16) Acetone	3.78	43	17594	15.914	ug/l	96
18) Methyl Acetate	4.29	43	5865	2.594	ug/l #	89
20) Methylene Chloride	4.50	84	29650	11.187	ug/l	97
43) Isopropyl Acetate	8.13	43	54022	10.082	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062620\
Data File : VN062173.D
Acq On : 27 Jun 2020 5:29
Operator : JC/MD
Sample : PB129754ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129754ZHE#04

Quant Time: Jun 27 07:09:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

