


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062620\
Data File : VN062127.D
Acq On : 26 Jun 2020 8:45
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Title : SW846 8260
Last Update : Wed Jun 24 15:33:00 2020



AutoFind: Scans 3296, 3297, 3298; Background Corrected with Scan 3283

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	9940	PASS
75	95	30	60	49.3	25386	PASS
95	95	100	100	100.0	51501	PASS
96	95	5	9	6.8	3509	PASS
173	174	0.00	2	1.4	557	PASS
174	95	50	100	79.6	41013	PASS
175	174	5	9	7.5	3068	PASS
176	174	95	101	98.3	40328	PASS
177	176	5	9	6.0	2413	PASS