

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062623\
 Data File : VN078403.D
 Acq On : 26 Jun 2023 15:36
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 27 02:17:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	434456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	800034	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	722836	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	220424	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	301832	52.106	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.220%	
35) Dibromofluoromethane	8.177	113	290277	57.384	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 114.760%	
50) Toluene-d8	10.571	98	928559	47.616	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.240%	
62) 4-Bromofluorobenzene	12.859	95	298873	49.299	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.600%	
Target Compounds						
					Qvalue	
85) sec-Butylbenzene	13.618	105	4151	0.216	ug/l	98
86) p-Isopropyltoluene	13.741	119	3208	0.206	ug/l	92
87) 1,3-Dichlorobenzene	13.741	146	2344	0.286	ug/l	84
89) n-Butylbenzene	14.059	91	5243	0.427	ug/l	89
91) 1,2-Dichlorobenzene	14.112	146	2270	0.284	ug/l	86
93) 1,2,4-Trichlorobenzene	15.406	180	2340	0.821	ug/l	95
95) Naphthalene	15.647	128	8741	0.912	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062623\
Data File : VN078403.D
Acq On : 26 Jun 2023 15:36
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 27 02:17:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

