

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062623\
 Data File : VN078405.D
 Acq On : 26 Jun 2023 16:25
 Operator : JC\MD
 Sample : 03376-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-062123

Quant Time: Jun 27 02:18:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	430474	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	783894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	726329	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	238996	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	300128	52.291	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.580%
35) Dibromofluoromethane	8.171	113	278993	56.289	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.580%
50) Toluene-d8	10.577	98	919407	48.117	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.240%
62) 4-Bromofluorobenzene	12.859	95	300157	50.531	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062623\
Data File : VN078405.D
Acq On : 26 Jun 2023 16:25
Operator : JC\MD
Sample : 03376-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-062123

Quant Time: Jun 27 02:18:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

