

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062623\
 Data File : VN078406.D
 Acq On : 26 Jun 2023 16:49
 Operator : JC\MD
 Sample : 03376-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB-GW-062123

Quant Time: Jun 27 02:18:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	443419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	828258	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	754234	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	235103	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	307689	52.044	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.080%
35) Dibromofluoromethane	8.171	113	296923	56.698	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	113.400%
50) Toluene-d8	10.576	98	954017	47.254	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.500%
62) 4-Bromofluorobenzene	12.859	95	310137	49.414	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062623\
Data File : VN078406.D
Acq On : 26 Jun 2023 16:49
Operator : JC\MD
Sample : 03376-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB-GW-062123

Quant Time: Jun 27 02:18:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

