

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062623\
 Data File : VN078407.D
 Acq On : 26 Jun 2023 17:13
 Operator : JC\MD
 Sample : 03417-02 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1621

Quant Time: Jun 27 02:18:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	413860	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	766920	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	675993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	195213	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	289516	52.467	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.940%	
35) Dibromofluoromethane	8.177	113	278014	57.333	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 114.660%	
50) Toluene-d8	10.576	98	882117	47.187	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 94.380%	
62) 4-Bromofluorobenzene	12.859	95	275347	47.380	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.760%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	133503	84.055	ug/l	94
52) Toluene	10.635	92	8584	0.655	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062623\
Data File : VN078407.D
Acq On : 26 Jun 2023 17:13
Operator : JC\MD
Sample : 03417-02 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1621

Quant Time: Jun 27 02:18:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

