

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062625\
 Data File : VN087188.D
 Acq On : 26 Jun 2025 12:44
 Operator : JC\MD
 Sample : Q2388-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-12

Quant Time: Jun 27 01:24:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	145145	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	283245	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	255868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	113541	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	103497	53.258	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.520%
35) Dibromofluoromethane	8.177	113	83635	49.825	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.640%
50) Toluene-d8	10.565	98	306992	46.199	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.400%
62) 4-Bromofluorobenzene	12.847	95	112916	45.737	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.480%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	33557	32.562	ug/l	96
18) Methyl Acetate	5.053	43	26052	8.675	ug/l	98
20) Methylene Chloride	5.312	84	5381	2.789	ug/l	97
37) Ethyl Acetate	7.582	43	15909	4.996	ug/l	97
43) Isopropyl Acetate	8.694	43	10889	2.131	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062625\
Data File : VN087188.D
Acq On : 26 Jun 2025 12:44
Operator : JC\MD
Sample : Q2388-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-12

Quant Time: Jun 27 01:24:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

