

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062625\
 Data File : VN087189.D
 Acq On : 26 Jun 2025 13:05
 Operator : JC\MD
 Sample : Q2389-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-J-I

Quant Time: Jun 27 01:24:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	147451	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	264397	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	246322	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	114926	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	106912	54.154	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.300%
35) Dibromofluoromethane	8.171	113	86053	54.920	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.840%
50) Toluene-d8	10.565	98	298690	48.154	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.300%
62) 4-Bromofluorobenzene	12.847	95	114464	49.669	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.340%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	31694	30.273	ug/l	89
18) Methyl Acetate	5.053	43	26427	8.662	ug/l	95
20) Methylene Chloride	5.300	84	5550	2.832	ug/l	89
37) Ethyl Acetate	7.571	43	17046	5.735	ug/l	96
43) Isopropyl Acetate	8.700	43	11293	2.367	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062625\
Data File : VN087189.D
Acq On : 26 Jun 2025 13:05
Operator : JC\MD
Sample : Q2389-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-J-I

Quant Time: Jun 27 01:24:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

