

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062625\
 Data File : VN087194.D
 Acq On : 26 Jun 2025 14:49
 Operator : JC\MD
 Sample : Q2405-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-M/N

Quant Time: Jun 27 01:26:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	178724	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	364399	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	342365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	158203	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	142307	59.470	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.940%
35) Dibromofluoromethane	8.171	113	110655	51.241	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.480%
50) Toluene-d8	10.565	98	445997	52.170	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.340%
62) 4-Bromofluorobenzene	12.847	95	154950	48.785	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.560%
Target Compounds						
						Qvalue
16) Acetone	4.453	43	29680	23.389	ug/l #	87
18) Methyl Acetate	5.047	43	24836	6.716	ug/l	91
20) Methylene Chloride	5.312	84	7208	3.034	ug/l	95
37) Ethyl Acetate	7.571	43	17374	4.241	ug/l	97
43) Isopropyl Acetate	8.694	43	10385	1.580	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062625\
Data File : VN087194.D
Acq On : 26 Jun 2025 14:49
Operator : JC\MD
Sample : Q2405-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-M/N

Quant Time: Jun 27 01:26:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

