


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN062625\  
Data File : VN087180.D  
Acq On : 26 Jun 2025 08:40  
Operator : JC\MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N062625W.M  
Title : SW846 8260  
Last Update : Fri Jun 27 05:55:21 2025



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1844

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw   | Result    |
|--------|---------|--------|--------|-------|-------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn   | Pass/Fail |
| 50     | 95      | 15     | 40     | 19.4  | 3438  | PASS      |
| 75     | 95      | 30     | 60     | 49.5  | 8798  | PASS      |
| 95     | 95      | 100    | 100    | 100.0 | 17765 | PASS      |
| 96     | 95      | 5      | 9      | 7.0   | 1235  | PASS      |
| 173    | 174     | 0.00   | 2      | 0.9   | 115   | PASS      |
| 174    | 95      | 50     | 100    | 70.8  | 12584 | PASS      |
| 175    | 174     | 5      | 9      | 7.3   | 922   | PASS      |
| 176    | 174     | 95     | 101    | 97.1  | 12215 | PASS      |
| 177    | 176     | 5      | 9      | 7.8   | 956   | PASS      |

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z   | abund. | m/z    | abund. |
|-------|--------|-------|--------|-------|--------|--------|--------|
| 35.90 | 57     | 56.00 | 231    | 71.80 | 119    | 91.95  | 431    |
| 36.10 | 73     | 56.95 | 514    | 72.80 | 188    | 93.00  | 741    |
| 37.00 | 849    | 60.10 | 73     | 73.00 | 627    | 94.00  | 213    |
| 37.95 | 781    | 61.00 | 922    | 74.00 | 2999   | 95.00  | 1776   |
| 39.05 | 371    | 62.00 | 975    | 75.00 | 8798   | 95.95  | 1235   |
| 39.85 | 123    | 62.95 | 698    | 76.00 | 819    | 142.90 | 53     |
| 44.95 | 165    | 63.90 | 68     | 77.00 | 102    | 171.95 | 161    |
| 46.95 | 278    | 64.90 | 59     | 78.85 | 556    | 172.85 | 115    |
| 49.00 | 699    | 68.00 | 1682   | 80.95 | 553    | 173.95 | 12584  |
| 50.00 | 3438   | 68.95 | 1706   | 86.95 | 943    | 174.95 | 922    |
| 51.05 | 1031   | 69.90 | 69     | 87.95 | 826    | 175.90 | 12215  |

Average of 12.841 to 12.853 min.: VN087180.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z | abund. | m/z | abund. | m/z | abund. |
|--------|--------|-----|--------|-----|--------|-----|--------|
| 176.90 | 956    |     |        |     |        |     |        |
| 207.00 | 52     |     |        |     |        |     |        |

Instrument :

MSVOA\_N

ClientSampleId :

BFB