

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062718\
 Data File : VN049568.D
 Acq On : 27 Jun 2018 13:31
 Operator : MD\SY
 Sample : VN0627WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jun 27 17:22:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1868515	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	2902163	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2415334	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	978719	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	1157677	46.71	ug/l	0.00
Spiked Amount			Recovery	=	93.42%	
35) Dibromofluoromethane	7.59	113	1025831	43.69	ug/l	0.00
Spiked Amount			Recovery	=	87.38%	
50) Toluene-d8	10.09	98	3935447	44.02	ug/l	0.00
Spiked Amount			Recovery	=	88.04%	
62) 4-Bromofluorobenzene	12.40	95	1161540	39.29	ug/l	0.00
Spiked Amount			Recovery	=	78.58%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.80	59	515	0.35	ug/l #	70
13) Acrolein	3.60	56	3073	2.01	ug/l #	44
16) Acetone	3.82	43	214962	35.72	ug/l	91
17) Carbon Disulfide	4.05	76	16614	0.25	ug/l	100
18) Methyl Acetate	4.33	43	96910	5.28	ug/l	96
20) Methylene Chloride	4.55	84	11980	0.48	ug/l #	82
25) 2-Butanone	6.84	43	4485	0.56	ug/l	92
29) Tetrahydrofuran	7.22	42	2694	0.52	ug/l #	66
31) Cyclohexane	7.66	56	30809	0.67	ug/l #	1
36) 1,1-Dichloropropene	7.66	75	124714	3.75	ug/l #	51
38) Carbon Tetrachloride	7.66	117	163903	4.76	ug/l #	17
43) Isopropyl Acetate	8.26	43	373504	10.81	ug/l #	64
49) 1,4-Dioxane	9.18	88	65	0.29	ug/l #	1
51) 4-Methyl-2-Pentanone	9.99	43	4206	0.23	ug/l #	87
58) 2-Chloroethyl Vinyl ether	9.69	63	64	14.33	ug/l #	51
59) 2-Hexanone	10.76	43	6009	0.51	ug/l	80
75) 1,1,2,2-Tetrachloroethane	12.50	83	1492	Below Cal	#	78
76) 1,2,3-Trichloropropane	12.40	75	556354	33.32	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	556354	105.30	ug/l #	9
87) 1,3-Dichlorobenzene	13.28	146	7770	0.21	ug/l	96
88) 1,4-Dichlorobenzene	13.28	146	7770	0.21	ug/l	97
89) n-Butylbenzene	13.62	91	10804	0.21	ug/l	96
90) Hexachloroethane	13.88	117	167	Below Cal	#	9
93) 1,2,4-Trichlorobenzene	14.91	180	7949	5.17	ug/l	90
94) Hexachlorobutadiene	15.01	225	5857	Below Cal		92
95) Naphthalene	15.14	128	19430	2.79	ug/l	99
96) 1,2,3-Trichlorobenzene	15.32	180	9293	3.98	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062718\
Data File : VN049568.D
Acq On : 27 Jun 2018 13:31
Operator : MD\SY
Sample : VN0627WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Jun 27 17:22:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.66 min Scan# 1889

Delta R.T. -0.00 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31

Instrument :

MSVOA_N


ClientSampled :

Tot Ion:168 Resp: 1868515

Ion Ratio Lower Upper

168 100

99 51.2 45.8 68.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.66

800000



600000

400000

200000

0

Time-->



#11

Tert butyl alcohol

Concen: 0.35 ug/l

RT: 4.80 min Scan# 998

Delta R.T. 0.00 min

Lab File: VN049568.D


Acq: 27 Jun 2018 13:31

Tgt Ion: 59 Resp: 515

Ion Ratio Lower Upper

59 100

57 0.0 9.0 13.6#



Abundance Ion 59.00 (58.70 to 59.70): VNO

Ion 57.00 (56.70 to 57.70): VNO

4.80

500

400


300


200

100

0

Time-->





#13

Acrolein

Concen: 2.01 ug/l

RT: 3.60 min Scan# 626

Delta R.T. -0.00 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 56 Resp: 3073

Ion Ratio Lower Upper




56 100

55 25.1 57.1 85.7#



Abundance Ion 56.00 (55.70 to 56.70): VN0

Ion 55.00 (54.70 to 55.70): VN0



#16

Acetone

Concen: 35.72 ug/l

RT: 3.82 min Scan# 694

Delta R.T. 0.00 min

Lab File: VN049568.D


Acq: 27 Jun 2018 13:31

Tgt Ion: 43 Resp: 214962

Ion Ratio Lower Upper



43 100


58 32.6 22.4 33.6



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0









#17
Carbon Disulfide
Concen: 0.25 ug/l
RT: 4.05 min Scan# 765
Delta R.T. -0.00 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 76 Resp: 16614
Ion Ratio Lower Upper
76 100
78 9.0 7.4 11.0



#18
Methyl Acetate
Concen: 5.28 ug/l
RT: 4.33 min Scan# 852
Delta R.T. 0.00 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31

Tgt Ion: 43 Resp: 96910
Ion Ratio Lower Upper
43 100
74 23.6 17.4 26.0









#20
Methvlene Chloride
Concen: 0.48 ug/l
RT: 4.55 min Scan# 922
Delta R.T. 0.00 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 84 Resp: 11980
Ion Ratio Lower Upper
84 100
49 103.8 106.9 160.3#
51 33.5 32.7 49.1
86 67.0 49.8 74.8



#25
2-Butanone
Concen: 0.56 ug/l
RT: 6.84 min Scan# 1633
Delta R.T. -0.00 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31

Tgt Ion: 43 Resp: 4485
Ion Ratio Lower Upper
43 100
72 21.3 20.2 30.2









#29
Tetrahydrofuran
Concen: 0.52 ug/l
RT: 7.22 min Scan# 1752
Delta R.T. 0.01 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31

Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 42 Resp: 2694
Ion Ratio Lower Upper
42 100
72 25.2 32.2 48.4#
71 11.1 30.5 45.7#




Abundance Ion 42.00 (41.70 to 42.70): VN0
Ion 72.00 (71.70 to 72.70): VN0
Ion 71.00 (70.70 to 71.70): VN0






#31
Cyclohexane
Concen: 0.67 ug/l
RT: 7.66 min Scan# 1888
Delta R.T. 0.01 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31

Tgt Ion: 56 Resp: 30809
Ion Ratio Lower Upper
56 100
69 201.0 25.0 37.4#
84 152.1 64.6 96.8#



Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0





#33

1,2-Dichloroethane-d4

Concen: 46.71 ug/l

RT: 8.03 min Scan# 2002

Delta R.T. -0.00 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31

Instrument :

MSVOA_N





ClientSampleId :

Tot Ion: 65 Resp: 1157677

Ion Ratio Lower Upper

65 100

67 53.5 0.0 103.8



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 8.58 min Scan# 2175

Delta R.T. -0.00 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31




Tgt Ion: 114 Resp: 2902163


Ion Ratio Lower Upper

114 100

63 19.7 0.0 44.4

88 15.0 0.0 31.4









#35
 Dibromofluoromethane
 Concen: 43.69 ug/l
 RT: 7.59 min Scan# 1865
 Delta R.T. -0.01 min
 Lab File: VN049568.D
 Acq: 27 Jun 2018 13:31




Instrument :
 MSVOA_N
 ClientSampled :


Tot Ion:113 Resp: 1025831
 Ion Ratio Lower Upper
 113 100
 111 102.4 82.4 123.6
 192 22.3 15.0 22.6



#36
 1,1-Dichloropropene
 Concen: 3.75 ug/l
 RT: 7.66 min Scan# 1889
 Delta R.T. -0.13 min
 Lab File: VN049568.D
 Acq: 27 Jun 2018 13:31

Tgt Ion: 75 Resp: 124714
 Ion Ratio Lower Upper
 75 100
 110 8.8 17.0 50.9#
 77 0.1 24.4 36.6#





#38

Carbon Tetrachloride

Concen: 4.76 ug/l

RT: 7.66 min Scan# 1889

Delta R.T. -0.11 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31

Instrument :

MSVOA_N

ClientSampled :


Tot Ion:117 Resp: 163903

Ion Ratio Lower Upper




117 100

119 5.4 75.6 113.4#

121 0.0 24.7 37.1#



Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 10.81 ug/l

RT: 8.26 min Scan# 2076

Delta R.T. 0.10 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31


Tgt Ion: 43 Resp: 373504

Ion Ratio Lower Upper



43 100


61 0.0 14.2 21.4#

87 0.0 9.4 14.0#



Abundance Ion 43.00 (42.70 to 43.70): VNO
Ion 61.00 (60.70 to 61.70): VNO
Ion 87.00 (86.70 to 87.70): VNO









#49
1,4-Dioxane
Concen: 0.29 ug/l
RT: 9.18 min Scan# 2361
Delta R.T. -0.02 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 88 Resp: 65
Ion Ratio Lower Upper
88 100
43 0.0 30.9 46.3#
58 424.6 58.9 88.3#




Abundance Ion 88.00 (87.70 to 88.70): VN0
Ion 43.00 (42.70 to 43.70): VN0
Ion 58.00 (57.70 to 58.70): VN0






#50
Toluene-d8
Concen: 44.02 ug/l
RT: 10.09 min Scan# 2644
Delta R.T. 0.00 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31

Tgt Ion: 98 Resp: 3935447
Ion Ratio Lower Upper
98 100
100 63.6 49.8 74.8



Abundance Ion 98.00 (97.70 to 98.70): VN0
Ion 100.00 (99.70 to 100.70): VN0









#51
4-Methyl-2-Pentanone
Concen: 0.23 ug/l
RT: 9.99 min Scan# 2613
Delta R.T. 0.00 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 4206
Ion Ratio Lower Upper
43 100
58 44.1 29.0 43.6#



#58
2-Chloroethyl Vinyl ether
Concen: 14.33 ug/l
RT: 9.69 min Scan# 2519
Delta R.T. -0.01 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31

Tgt Ion: 63 Resp: 64
Ion Ratio Lower Upper
63 100
106 0.0 19.4 29.2#





#59

2-Hexanone

Concen: 0.51 ug/l

RT: 10.76 min Scan# 2852

Delta R.T. 0.01 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 43 Resp: 6009

Ion Ratio Lower Upper




43 100

58 64.4 25.2 75.6



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#62

4-Bromofluorobenzene

Concen: 39.29 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.00 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31


Tgt Ion: 95 Resp: 1161540

Ion Ratio Lower Upper

95 100

174 89.3 0.0 165.6



176 87.1 0.0 157.4




Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 3053
Delta R.T. -0.00 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion:117 Resp: 2415334
Ion Ratio Lower Upper
117 100
82 53.3 45.9 68.9
119 32.0 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.35 min Scan# 3656
Delta R.T. -0.00 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31

Tgt Ion:152 Resp: 978719
Ion Ratio Lower Upper
152 100
115 56.4 28.7 86.1
150 157.0 0.0 351.0





#75

1,1,2,2-Tetrachloroethane

Concen: Below Cal

RT: 12.50 min Scan# 3394

Delta R.T. -0.00 min

Lab File: VN049568.D





Acq: 27 Jun 2018 13:31

Instrument :

MSVOA_N

ClientSampled :

Tot Ion	83	Resp	1492
Ion Ratio	Lower	Upper	
83	100		
131	0.0	5.0	15.0#
85	47.3	32.0	96.0



#76

1,2,3-Trichloropropane

Concen: 33.32 ug/l




RT: 12.40 min Scan# 3363


Delta R.T. -0.15 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31

Tgt Ion	75	Resp	556354
Ion Ratio	Lower	Upper	
75	100		
77	1.3	0.0	0.0#









#81
trans-1,4-Dichloro-2-butene
Concen: 105.30 ug/l
RT: 12.40 min Scan# 3363
Delta R.T. 0.10 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 75 Resp: 556354
Ion Ratio Lower Upper
75 100
53 0.2 82.7 124.1#
89 0.0 36.5 54.7#



#87
1,3-Dichlorobenzene
Concen: 0.21 ug/l
RT: 13.28 min Scan# 3636
Delta R.T. -0.00 min
Lab File: VN049568.D
Acq: 27 Jun 2018 13:31

Tgt Ion: 146 Resp: 7770
Ion Ratio Lower Upper
146 100
111 34.5 19.9 59.6
148 64.8 32.1 96.5









#88
 1,4-Dichlorobenzene
 Concen: 0.21 ug/l
 RT: 13.28 min Scan# 3636
 Delta R.T. -0.09 min
 Lab File: VN049568.D
 Acq: 27 Jun 2018 13:31




Instrument :
 MSVOA_N
 ClientSampled :


Tot Ion:146 Resp: 7770
 Ion Ratio Lower Upper
 146 100
 111 34.5 19.1 57.1
 148 64.8 32.0 96.2



#89
 n-Butylbenzene
 Concen: 0.21 ug/l
 RT: 13.62 min Scan# 3741
 Delta R.T. -0.00 min
 Lab File: VN049568.D
 Acq: 27 Jun 2018 13:31

Tgt Ion: 91 Resp: 10804
 Ion Ratio Lower Upper
 91 100
 92 53.1 25.1 75.4
 134 26.6 12.0 36.0





#90

Hexachloroethane

Concen: Below Cal

RT: 13.88 min Scan# 3823

Delta R.T. 0.01 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31

Instrument :

MSVOA_N


ClientSampled :

Tot Ion:117 Resp: 167

Ion Ratio Lower Upper




117 100

201 0.0 40.1 120.2#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V



#93

1,2,4-Trichlorobenzene

Concen: 5.17 ug/l

RT: 14.91 min Scan# 4144

Delta R.T. 0.00 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31


Tgt Ion:180 Resp: 7949

Ion Ratio Lower Upper

180 100

182 105.2 46.9 140.8



145 31.3 14.7 44.1




Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.80 (144.50 to 145.50): V





#94

Hexachlorobutadiene

Concen: Below Cal

RT: 15.01 min Scan# 4175

Delta R.T. 0.00 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31


Instrument :

MSVOA_N

ClientSampled :

Tot Ion:225 Resp: 5857




Ion	Ratio	Lower	Upper
225	100		
223	75.6	32.1	96.3
227	65.1	31.8	95.3



Abundance Ion 224.70 (224.40 to 225.40): V

Ion 222.70 (222.40 to 223.40): V

Ion 226.70 (226.40 to 227.40): V



#95

Naphthalene

Concen: 2.79 ug/l

RT: 15.14 min Scan# 4214


Delta R.T. 0.00 min

Lab File: VN049568.D

Acq: 27 Jun 2018 13:31

Tgt Ion:128 Resp: 19430



Ion	Ratio	Lower	Upper
128	100		
127	13.5	10.6	16.0
129	11.2	8.5	12.7




Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V





#96
 1,2,3-Trichlorobenzene
 Concen: 3.98 ug/l
 RT: 15.32 min Scan# 4270
 Delta R.T. 0.00 min
 Lab File: VN049568.D
 Acq: 27 Jun 2018 13:31

Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
180	100		
182	90.3	47.3	141.8
145	31.0	15.6	46.7

