

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062718\
 Data File : VN049578.D
 Acq On : 27 Jun 2018 17:48
 Operator : MD\SY
 Sample : PB110627ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jun 28 01:24:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1098635	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1761983	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1382834	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	451739	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	717791	49.26	ug/l	0.00
Spiked Amount			Recovery	=	98.52%	
35) Dibromofluoromethane	7.59	113	600015	42.09	ug/l	0.00
Spiked Amount			Recovery	=	84.18%	
50) Toluene-d8	10.09	98	2328218	42.89	ug/l	0.00
Spiked Amount			Recovery	=	85.78%	
62) 4-Bromofluorobenzene	12.40	95	603895	33.65	ug/l	0.00
Spiked Amount			Recovery	=	67.30%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.80	59	322	0.38	ug/l #	70
13) Acrolein	3.61	56	712	0.79	ug/l #	1
16) Acetone	3.82	43	92422	26.12	ug/l	91
18) Methyl Acetate	4.33	43	186875	17.32	ug/l	94
20) Methylene Chloride	4.55	84	34825	2.36	ug/l #	87
25) 2-Butanone	6.85	43	9457	1.99	ug/l	98
29) Tetrahydrofuran	7.22	42	1401	0.46	ug/l #	60
31) Cyclohexane	7.66	56	19304	0.72	ug/l #	1
36) 1,1-Dichloropropene	7.66	75	74632	3.70	ug/l #	51
38) Carbon Tetrachloride	7.67	117	99223	4.75	ug/l #	17
43) Isopropyl Acetate	8.27	43	415117	19.79	ug/l #	64
49) 1,4-Dioxane	9.22	88	30	0.22	ug/l #	1
51) 4-Methyl-2-Pentanone	9.99	43	8139	0.74	ug/l	88
58) 2-Chloroethyl Vinyl ether	9.67	63	75	14.34	ug/l #	51
59) 2-Hexanone	10.77	43	1587	0.22	ug/l	76
65) Chlorobenzene	11.44	112	19263	0.54	ug/l	98
71) Bromoform	12.40	173	3515	0.41	ug/l #	1
75) 1,1,2,2-Tetrachloroethane	12.50	83	107	Below	Cal #	26
76) 1,2,3-Trichloropropane	12.40	75	292926	38.00	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	292926	120.11	ug/l #	9
90) Hexachloroethane	14.00	117	108	Below	Cal #	9
92) 1,2-Dibromo-3-Chloropropan	14.25	75	535	0.39	ug/l #	4
93) 1,2,4-Trichlorobenzene	14.92	180	252	4.77	ug/l #	26
95) Naphthalene	15.13	128	982	2.16	ug/l #	53
96) 1,2,3-Trichlorobenzene	15.33	180	240	3.49	ug/l	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062718\
Data File : VN049578.D
Acq On : 27 Jun 2018 17:48
Operator : MD\SY
Sample : PB110627ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Jun 28 01:24:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. 0.00 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion:168 Resp: 1098635

Ion Ratio Lower Upper

168 100

99 51.0 45.8 68.6



Abundance Ion 167.90 (167.60 to 168.60): VN049263.D (-1869) (-)

Ion 98.90 (98.60 to 99.60): VN049263.D (-1869) (-)

7.67

400000



300000

200000

100000

0

Time-->



#11

Tert butyl alcohol

Concen: 0.38 ug/l

RT: 4.80 min Scan# 997

Delta R.T. 0.00 min

Lab File: VN049578.D


Acq: 27 Jun 2018 17:48

Tgt Ion: 59 Resp: 322

Ion Ratio Lower Upper

59 100

57 0.0 9.0 13.6#



Abundance Ion 59.00 (58.70 to 59.70): VN049263.D (-966) (-)

Ion 57.00 (56.70 to 57.70): VN049263.D (-966) (-)

4.80

400


300


200

100

0

Time-->





#13

Acrolein

Concen: 0.79 ug/l

RT: 3.61 min Scan# 627

Delta R.T. 0.00 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48

Instrument :

MSVOA_N





ClientSampled :

Tot Ion: 56 Resp: 712

Ion Ratio Lower Upper

56 100

55 226.3 57.1 85.7#



#16

Acetone

Concen: 26.12 ug/l

RT: 3.82 min Scan# 694

Delta R.T. 0.00 min

Lab File: VN049578.D




Acq: 27 Jun 2018 17:48


Tgt Ion: 43 Resp: 92422

Ion Ratio Lower Upper

43 100

58 32.9 22.4 33.6









#18
Methvl Acetate
Concen: 17.32 ug/l
RT: 4.33 min Scan# 851
Delta R.T. 0.00 min
Lab File: VN049578.D
Acq: 27 Jun 2018 17:48




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 43 Resp: 186875
Ion Ratio Lower Upper
43 100
74 24.6 17.4 26.0



#20
Methylene Chloride
Concen: 2.36 ug/l
RT: 4.55 min Scan# 921
Delta R.T. 0.00 min
Lab File: VN049578.D
Acq: 27 Jun 2018 17:48

Tgt Ion: 84 Resp: 34825
Ion Ratio Lower Upper
84 100
49 115.5 106.9 160.3
51 32.5 32.7 49.1#
86 67.3 49.8 74.8









#25
2-Butanone
Concen: 1.99 ug/l
RT: 6.85 min Scan# 1635
Delta R.T. 0.01 min
Lab File: VN049578.D
Acq: 27 Jun 2018 17:48




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 43 Resp: 9457
Ion Ratio Lower Upper
43 100
72 24.0 20.2 30.2



#29
Tetrahydrofuran
Concen: 0.46 ug/l
RT: 7.22 min Scan# 1751
Delta R.T. 0.00 min
Lab File: VN049578.D
Acq: 27 Jun 2018 17:48

Tgt Ion: 42 Resp: 1401
Ion Ratio Lower Upper
42 100
72 16.6 32.2 48.4#
71 13.3 30.5 45.7#









#31
Cyclohexane
Concen: 0.72 ug/l
RT: 7.66 min Scan# 1889
Delta R.T. 0.01 min
Lab File: VN049578.D
Acq: 27 Jun 2018 17:48

Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 56 Resp: 19304
Ion Ratio Lower Upper
56 100
69 171.8 25.0 37.4#
84 139.1 64.6 96.8#




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 49.26 ug/l
RT: 8.03 min Scan# 2002
Delta R.T. 0.00 min
Lab File: VN049578.D
Acq: 27 Jun 2018 17:48

Tgt Ion: 65 Resp: 717791
Ion Ratio Lower Upper
65 100
67 53.6 0.0 103.8



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.59 min Scan# 2176
 Delta R.T. 0.00 min
 Lab File: VN049578.D
 Acq: 27 Jun 2018 17:48




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion:114 Resp: 1761983
 Ion Ratio Lower Upper
 114 100
 63 19.7 0.0 44.4
 88 15.4 0.0 31.4



#35
 Dibromofluoromethane
 Concen: 42.09 ug/l
 RT: 7.59 min Scan# 1867
 Delta R.T. 0.00 min
 Lab File: VN049578.D
 Acq: 27 Jun 2018 17:48

Tgt Ion:113 Resp: 600015
 Ion Ratio Lower Upper
 113 100
 111 101.7 82.4 123.6
 192 22.3 15.0 22.6









#36
 1,1-Dichloropropene
 Concen: 3.70 ug/l
 RT: 7.66 min Scan# 1889
 Delta R.T. -0.13 min
 Lab File: VN049578.D
 Acq: 27 Jun 2018 17:48

Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion: 75 Resp: 74632
 Ion Ratio Lower Upper
 75 100
 110 9.2 17.0 50.9#
 77 0.3 24.4 36.6#




Abundance Ion 74.90 (74.60 to 75.60): VN049263.D (-1905) (-)
 Ion 109.90 (109.60 to 110.60): VN049263.D (-1905) (-)
 Ion 76.90 (76.60 to 77.60): VN049263.D (-1905) (-)






#38
 Carbon Tetrachloride
 Concen: 4.75 ug/l
 RT: 7.67 min Scan# 1890
 Delta R.T. -0.11 min
 Lab File: VN049578.D
 Acq: 27 Jun 2018 17:48

Tgt Ion: 117 Resp: 99223
 Ion Ratio Lower Upper
 117 100
 119 5.7 75.6 113.4#
 121 0.0 24.7 37.1#



Abundance Ion 116.80 (116.50 to 117.50): VN049263.D (-1904) (-)
 Ion 118.80 (118.50 to 119.50): VN049263.D (-1904) (-)
 Ion 120.80 (120.50 to 121.50): VN049263.D (-1904) (-)





#43

Isopropyl Acetate

Concen: 19.79 ug/l

RT: 8.27 min Scan# 2079

Delta R.T. 0.11 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48

Instrument :

MSVOA_N

ClientSampled :


Tot Ion: 43 Resp: 415117

Ion Ratio Lower Upper

43 100

61 0.0 14.2 21.4#




87 0.0 9.4 14.0#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#49

1,4-Dioxane

Concen: 0.22 ug/l

RT: 9.22 min Scan# 2373

Delta R.T. 0.02 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48


Tgt Ion: 88 Resp: 30

Ion Ratio Lower Upper

88 100

43 0.0 30.9 46.3#



58 240.0 58.9 88.3#




Abundance Ion 88.00 (87.70 to 88.70): VN0

Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0









#50
Toluene-d8
Concen: 42.89 ug/l
RT: 10.09 min Scan# 2644
Delta R.T. 0.00 min
Lab File: VN049578.D
Acq: 27 Jun 2018 17:48




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 98 Resp: 2328218
Ion Ratio Lower Upper
98 100
100 63.7 49.8 74.8



#51
4-Methyl-2-Pentanone
Concen: 0.74 ug/l
RT: 9.99 min Scan# 2611
Delta R.T. -0.00 min
Lab File: VN049578.D
Acq: 27 Jun 2018 17:48

Tgt Ion: 43 Resp: 8139
Ion Ratio Lower Upper
43 100
58 29.2 29.0 43.6





#58

2-Chloroethyl Vinyl ether

Concen: 14.34 ug/l

RT: 9.67 min Scan# 2512

Delta R.T. -0.03 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48


Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 63 Resp: 75




Ion Ratio Lower Upper

63 100

106 0.0 19.4 29.2#



Abundance Ion 62.90 (62.60 to 63.60): VN049578.D (-2512) (-)



#59

2-Hexanone

Concen: 0.22 ug/l

RT: 10.77 min Scan# 2854

Delta R.T. 0.01 min

Lab File: VN049578.D


Acq: 27 Jun 2018 17:48

Tgt Ion: 43 Resp: 1587



Ion Ratio Lower Upper


43 100

58 33.7 25.2 75.6



Abundance Ion 43.00 (42.70 to 43.70): VN049578.D (-2854) (-)









#62
4-Bromofluorobenzene
Concen: 33.65 ug/l
RT: 12.40 min Scan# 3363
Delta R.T. 0.00 min
Lab File: VN049578.D
Acq: 27 Jun 2018 17:48




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion	95	Resp	603895
Ion Ratio	Lower	Upper	
95	100		
174	90.5	0.0	165.6
176	86.6	0.0	157.4



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 3054
Delta R.T. 0.00 min
Lab File: VN049578.D
Acq: 27 Jun 2018 17:48

Tgt Ion	117	Resp	1382834
Ion Ratio	Lower	Upper	
117	100		
82	53.7	45.9	68.9
119	31.3	25.5	38.3





#65

Chlorobenzene

Concen: 0.54 ug/l

RT: 11.44 min Scan# 3063

Delta R.T. 0.00 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48

Instrument :

MSVOA_N


ClientSampled :

Tot Ion:112 Resp: 19263

Ion Ratio Lower Upper




112 100

114 33.8 26.2 39.4



Abundance Ion 111.90 (111.60 to 112.60): V

Ion 113.90 (113.60 to 114.60): V



#71

Bromoform

Concen: 0.41 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.28 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48


Tgt Ion:173 Resp: 3515

Ion Ratio Lower Upper

173 100

175 906.5 24.5 73.5#



254 0.0 0.1 0.1#




Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 13.35 min Scan# 3656

Delta R.T. 0.00 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48

Instrument :

MSVOA_N

ClientSampled :


Tot Ion:152 Resp: 451739

Ion Ratio Lower Upper

152 100

115 56.5 28.7 86.1




150 156.4 0.0 351.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#75

1,1,2,2-Tetrachloroethane

Concen: Below Cal

RT: 12.50 min Scan# 3392

Delta R.T. -0.01 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48


Tgt Ion: 83 Resp: 107

Ion Ratio Lower Upper

83 100

131 0.0 5.0 15.0#



85 0.0 32.0 96.0#




Abundance Ion 82.90 (82.60 to 83.60): V

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): V





#76

1,2,3-Trichloropropane

Concen: 38.00 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.15 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion: 75 Resp: 292926

Ion Ratio Lower Upper




75 100

77 1.3 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VNO

Ion 77.00 (76.70 to 77.70): VNO



#81

trans-1,4-Dichloro-2-butene

Concen: 120.11 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.10 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48


Tgt Ion: 75 Resp: 292926

Ion Ratio Lower Upper

75 100

53 0.3 82.7 124.1#



89 0.0 36.5 54.7#




Abundance Ion 74.90 (74.60 to 75.60): VNO

Ion 53.00 (52.70 to 53.70): VNO

Ion 88.90 (88.60 to 89.60): VNO





#90

Hexachloroethane

Concen: Below Cal

RT: 14.00 min Scan# 3859

Delta R.T. 0.12 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48

Instrument :

MSVOA_N


ClientSampled :

Tot Ion:117 Resp: 108

Ion Ratio Lower Upper




117 100

201 0.0 40.1 120.2#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 0.39 ug/l

RT: 14.25 min Scan# 3938

Delta R.T. -0.02 min

Lab File: VN049578.D

Acq: 27 Jun 2018 17:48


Tgt Ion: 75 Resp: 535

Ion Ratio Lower Upper

75 100

155 0.0 38.8 116.3#



157 0.0 51.0 153.1#




Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V









#93
 1,2,4-Trichlorobenzene
 Concen: 4.77 ug/l
 RT: 14.92 min Scan# 4145
 Delta R.T. 0.00 min
 Lab File: VN049578.D
 Acq: 27 Jun 2018 17:48

Instrument :
 MSVOA_N
 ClientSampled :

Tot Ion:180 Resp: 252
 Ion Ratio Lower Upper
 180 100
 182 170.6 46.9 140.8#
 145 0.0 14.7 44.1#




Abundance Ion 179.80 (179.50 to 180.50): V
 Ion 181.80 (181.50 to 182.50): V
 Ion 144.80 (144.50 to 145.50): V






#95
 Naphthalene
 Concen: 2.16 ug/l
 RT: 15.13 min Scan# 4211
 Delta R.T. -0.01 min
 Lab File: VN049578.D
 Acq: 27 Jun 2018 17:48

Tgt Ion:128 Resp: 982
 Ion Ratio Lower Upper
 128 100
 127 3.6 10.6 16.0#
 129 39.0 8.5 12.7#



Abundance Ion 128.00 (127.70 to 128.70): V
 Ion 127.00 (126.70 to 127.70): V
 Ion 129.00 (128.70 to 129.70): V





#96
 1,2,3-Trichlorobenzene
 Concen: 3.49 ug/l
 RT: 15.33 min Scan# 4272
 Delta R.T. 0.01 min
 Lab File: VN049578.D
 Acq: 27 Jun 2018 17:48

Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
180	100		
182	110.4	47.3	141.8
145	27.1	15.6	46.7

