

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062718\
 Data File : VN049568.D
 Acq On : 27 Jun 2018 13:31
 Operator : MD\SY
 Sample : VN0627WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0627WBL01

Quant Time: Jun 28 03:30:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1868515	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	2902163	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2415334	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	978719	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	1157677	46.71	ug/l	0.00
Spiked Amount						
			Recovery	=	93.42%	
35) Dibromofluoromethane	7.59	113	1025831	43.69	ug/l	0.00
Spiked Amount						
			Recovery	=	87.38%	
50) Toluene-d8	10.09	98	3935447	44.02	ug/l	0.00
Spiked Amount						
			Recovery	=	88.04%	
62) 4-Bromofluorobenzene	12.40	95	1161540	39.29	ug/l	0.00
Spiked Amount						
			Recovery	=	78.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	214962	35.72	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062718\
Data File : VN049568.D
Acq On : 27 Jun 2018 13:31
Operator : MD\SY
Sample : VN0627WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0627WBL01

Quant Time: Jun 28 03:30:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

