

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062718\
 Data File : VN049569.D
 Acq On : 27 Jun 2018 13:57
 Operator : MD\SY
 Sample : J3667-02 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CS-PAINT-CHIP

Quant Time: Jun 28 01:39:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1228814	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1994512	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1631196	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	712230	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	817269	50.14	ug/l	0.00
Spiked Amount			Recovery	=	100.28%	
35) Dibromofluoromethane	7.59	113	381632	23.65	ug/l	0.00
Spiked Amount			Recovery	=	47.30%	
50) Toluene-d8	10.09	98	2649505	43.12	ug/l	0.00
Spiked Amount			Recovery	=	86.24%	
62) 4-Bromofluorobenzene	12.40	95	813089	40.02	ug/l	0.00
Spiked Amount			Recovery	=	80.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.85	43	562389	142.12	ug/l	94
18) Methyl Acetate	4.34	43	122188	10.13	ug/l	95
20) Methylene Chloride	4.55	84	52571	3.18	ug/l	91
25) 2-Butanone	6.86	43	174174	32.78	ug/l	96
51) 4-Methyl-2-Pentanone	9.99	43	27212	2.19	ug/l	96
59) 2-Hexanone	10.75	43	91348	11.39	ug/l	95
95) Naphthalene	15.14	128	42400	4.22	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062718\
Data File : VN049569.D
Acq On : 27 Jun 2018 13:57
Operator : MD\SY
Sample : J3667-02 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CS-PAINT-CHIP

Quant Time: Jun 28 01:39:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

