

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062718\
 Data File : VN049576.D
 Acq On : 27 Jun 2018 16:57
 Operator : MD\SY
 Sample : PB110627ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110627ZHE#04

Quant Time: Jun 28 02:05:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1086691	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1782695	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1450724	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	550931	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	736553	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
35) Dibromofluoromethane	7.59	113	605841	42.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.02%	
50) Toluene-d8	10.09	98	2372950	43.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.42%	
62) 4-Bromofluorobenzene	12.40	95	664850	36.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.22%	
Target Compounds						
16) Acetone	3.82	43	90862	25.96	ug/l	93
18) Methyl Acetate	4.33	43	200031	18.75	ug/l	95
20) Methylene Chloride	4.55	84	48563	3.32	ug/l	84
25) 2-Butanone	6.85	43	9439	2.01	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062718\
Data File : VN049576.D
Acq On : 27 Jun 2018 16:57
Operator : MD\SY
Sample : PB110627ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110627ZHE#04

Quant Time: Jun 28 02:05:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

