


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062718\
 Data File : VN049566.D
 Acq On : 27 Jun 2018 9:32
 Operator : MD\SY
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Title : SW846 8260
 Last Update : Sat Jun 16 01:02:01 2018



AutoFind: Scans 3362, 3363, 3364; Background Corrected with Scan 3347

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	31408	PASS
75	95	30	60	46.8	84130	PASS
95	95	100	100	100.0	179712	PASS
96	95	5	9	6.4	11477	PASS
173	174	0.00	2	1.2	1865	PASS
174	95	50	100	90.0	161685	PASS
175	174	5	9	7.3	11862	PASS
176	174	95	101	96.9	156714	PASS
177	176	5	9	6.0	9479	PASS