

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062719\
 Data File : VN056517.D
 Acq On : 27 Jun 2019 16:09
 Operator : JC/SP
 Sample : VSSTDICC008
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC008

Quant Time: Jun 28 05:55:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	55548	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	95574	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	58775	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	9513	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	51943	7.63	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	152.60%#
5) Dibromofluoromethane	7.59	113	39012	7.40	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	148.00%#
6) Toluene-d8	10.09	98	90537	7.59	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	151.80%#
7) 4-Bromofluorobenzene	12.39	95	24151	7.92	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	158.40%#

Target Compounds


	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.17	62	43576	7.847	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062719\
Data File : VN056517.D
Acq On : 27 Jun 2019 16:09
Operator : JC/SP
Sample : VSSTDICC008
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC008

Quant Time: Jun 28 05:55:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 03:22:32 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 5.000 ug/l

RT: 7.65 min Scan# 1243

Delta R.T. 0.00 min

Lab File: VN056517.D

Acq: 27 Jun 2019 16:09

Instrument :

MSVOA_N

ClientSampleId :


VSSTDICC008

Tot Ion:168 Resp: 55548




Ion Ratio Lower Upper

168 100

99 51.5 41.9 62.9



Abundance Ion 168.00 (167.70 to 168.70): VN056517.D (-1225) (-)



#2

Vinyl Chloride

Concen: 7.847 ug/l

RT: 2.17 min Scan# 150

Delta R.T. 0.00 min

Lab File: VN056517.D


Acq: 27 Jun 2019 16:09

Tgt Ion: 62 Resp: 43576



Ion Ratio Lower Upper

62 100

64 32.6 26.1 39.1



Abundance Ion 62.00 (61.70 to 62.70): VN056517.D (-98) (-)





#3

1,2-Dichloroethane-d4

Concen: 7.634 ug/l

RT: 8.01 min Scan# 1264

Delta R.T. 0.00 min

Lab File: VN056517.D

Acq: 27 Jun 2019 16:09

Instrument :

MSVOA_N

ClientSampleId :


VSSTDICC008

Tot Ion: 65 Resp: 51943

Ion Ratio Lower Upper

65 100

67 54.3 0.0 108.8



Abundance Ion 65.00 (64.70 to 65.70): VN0

Ion 67.00 (66.70 to 67.70): VN0

15000

10000



5000

0

8.01

7.80 8.00 8.20

Time-->



#4

1,4-Difluorobenzene

Concen: 5.000 ug/l

RT: 8.58 min Scan# 1297

Delta R.T. 0.00 min

Lab File: VN056517.D

Acq: 27 Jun 2019 16:09


Tgt Ion: 114 Resp: 95574

Ion Ratio Lower Upper

114 100

63 18.8 0.0 37.6

88 15.6 0.0 31.2



Abundance Ion 114.00 (113.70 to 114.70): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 88.00 (87.70 to 88.70): VN0

30000

25000

20000

15000

10000


5000

0

8.58

8.50 9.00

Time-->





#5

Dibromofluoromethane

Concen: 7.404 ug/l

RT: 7.59 min Scan# 1239

Delta R.T. 0.00 min

Lab File: VN056517.D

Acq: 27 Jun 2019 16:09

Instrument :

MSVOA_N

ClientSampleId :

VSSTDICC008


Tot Ion: 113 Resp: 39012

Ion Ratio Lower Upper

113 100

111 101.9 81.7 122.5




192 20.8 16.9 25.3



Abundance Ion 113.00 (112.70 to 113.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 192.00 (191.70 to 192.70): V



#6

Toluene-d8

Concen: 7.594 ug/l

RT: 10.09 min Scan# 1440

Delta R.T. 0.00 min

Lab File: VN056517.D


Acq: 27 Jun 2019 16:09

Tgt Ion: 98 Resp: 90537

Ion Ratio Lower Upper



98 100

100 63.0 49.7 74.5



Abundance Ion 98.00 (97.70 to 98.70): VN

Ion 100.00 (99.70 to 100.70): VN





#7

4-Bromofluorobenzene

Concen: 7.919 ug/l

RT: 12.39 min Scan# 1772

Delta R.T. -0.01 min

Lab File: VN056517.D

Acq: 27 Jun 2019 16:09

Instrument :

MSVOA_N

ClientSampleId :

VSSTDICC008


Tot Ion: 95 Resp: 24151

Ion Ratio Lower Upper

95 100

174 75.2 0.0 167.8




176 67.3 0.0 158.0



Abundance Ion 95.00 (94.70 to 95.70): VN056517.D (-1763) (-)

Ion 174.00 (173.70 to 174.70): VN056517.D (-1763) (-)

Ion 176.00 (175.70 to 176.70): VN056517.D (-1763) (-)



#8

Chlorobenzene-d5

Concen: 5.000 ug/l

RT: 11.41 min Scan# 1628

Delta R.T. 0.00 min

Lab File: VN056517.D

Acq: 27 Jun 2019 16:09


Tgt Ion: 117 Resp: 58775

Ion Ratio Lower Upper

117 100

82 53.2 43.0 64.4



119 31.0 24.7 37.1





Abundance Ion 117.00 (116.70 to 117.70): VN056517.D (-1617) (-)

Ion 82.00 (81.70 to 82.70): VN056517.D (-1617) (-)

Ion 119.00 (118.70 to 119.70): VN056517.D (-1617) (-)





#9

1,4-Dichlorobenzene-d4

Concen: 5.000 ug/l

RT: 13.34 min Scan# 1863

Delta R.T. -0.01 min

Lab File: VN056517.D

Acq: 27 Jun 2019 16:09

Instrument :

MSVOA_N

ClientSampleId :

VSSTDICC008

Tot Ion:152 Resp: 9513

Ion Ratio Lower Upper

152 100

115 56.8 27.0 80.8

150 159.5 0.0 306.6

Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V

