

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062719\
 Data File : VN056522.D
 Acq On : 27 Jun 2019 20:49
 Operator : JC/SP
 Sample : 0.05ppb LOD
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 0.05ppb LOD

Quant Time: Jun 28 03:30:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	52041	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	89398	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	55123	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	8935	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	30846	4.84	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	96.80%
5) Dibromofluoromethane	7.59	113	22866	4.64	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	92.80%
6) Toluene-d8	10.09	98	52717	4.73	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	94.60%
7) 4-Bromofluorobenzene	12.40	95	14227	4.99	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	99.80%


Target Compounds					Qvalue
2) Vinyl Chloride	2.17	62	251	0.048 ug/l #	45


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062719\
Data File : VN056522.D
Acq On : 27 Jun 2019 20:49
Operator : JC/SP
Sample : 0.05ppb LOD
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
0.05ppb LOD

Quant Time: Jun 28 03:30:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 03:22:32 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 5.000 ug/l

RT: 7.65 min Scan# 1243

Delta R.T. 0.00 min

Lab File: VN056522.D

Acq: 27 Jun 2019 20:49

Instrument :

MSVOA_N

ClientSampled :


0.05ppb LOD

Tot Ion:168 Resp: 52041

Ion Ratio Lower Upper

168 100

99 51.8 41.9 62.9



Abundance Ion 168.00 (167.70 to 168.70): VN056522.D (-1225) (-)

Ion 99.00 (98.70 to 99.70): VN056522.D (-1225) (-)

7.65

15000



10000

5000

0

7.40 7.60 7.80 8.00

Time-->



#2

Vinyl Chloride

Concen: 0.048 ug/l

RT: 2.17 min Scan# 150

Delta R.T. 0.00 min

Lab File: VN056522.D


Acq: 27 Jun 2019 20:49

Tgt Ion: 62 Resp: 251

Ion Ratio Lower Upper

62 100

64 63.6 26.1 39.1#



Abundance Ion 62.00 (61.70 to 62.70): VN056522.D (-98) (-)

Ion 64.00 (63.70 to 64.70): VN056522.D (-98) (-)

2.17

250

200

150


100


50

0

2.10 2.15 2.20 2.25

Time-->





#3

1,2-Dichloroethane-d4

Concen: 4.839 ug/l

RT: 8.01 min Scan# 1264

Delta R.T. 0.00 min

Lab File: VN056522.D

Acq: 27 Jun 2019 20:49

Instrument :

MSVOA_N

ClientSampleId :


0.05ppb LOD

Tot Ion: 65 Resp: 30846

Ion Ratio Lower Upper




65 100

67 54.3 0.0 108.8



Abundance Ion 65.00 (64.70 to 65.70): VN0

Ion 67.00 (66.70 to 67.70): VN0



#4

1,4-Difluorobenzene

Concen: 5.000 ug/l

RT: 8.58 min Scan# 1297

Delta R.T. 0.00 min

Lab File: VN056522.D

Acq: 27 Jun 2019 20:49


Tgt Ion: 114 Resp: 89398

Ion Ratio Lower Upper

114 100

63 18.8 0.0 37.6



88 15.6 0.0 31.2




Abundance Ion 114.00 (113.70 to 114.70): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 88.00 (87.70 to 88.70): VN0







#5
 Dibromofluoromethane
 Concen: 4.640 ug/l
 RT: 7.59 min Scan# 1239
 Delta R.T. 0.00 min
 Lab File: VN056522.D
 Acq: 27 Jun 2019 20:49

Instrument :
 MSVOA_N
 ClientSampleId :
 0.05ppb LOD



Tot Ion:113 Resp: 22866
 Ion Ratio Lower Upper
 113 100
 111 103.0 81.7 122.5
 192 21.0 16.9 25.3



Abundance Ion 113.00 (112.70 to 113.70): V
 Ion 111.00 (110.70 to 111.70): V
 Ion 192.00 (191.70 to 192.70): V




Time-->




#6
 Toluene-d8
 Concen: 4.727 ug/l
 RT: 10.09 min Scan# 1440
 Delta R.T. 0.00 min
 Lab File: VN056522.D
 Acq: 27 Jun 2019 20:49


Tgt Ion: 98 Resp: 52717
 Ion Ratio Lower Upper
 98 100
 100 62.4 49.7 74.5




Abundance Ion 98.00 (97.70 to 98.70): VN
 Ion 100.00 (99.70 to 100.70): VN



Time-->





#7

4-Bromofluorobenzene

Concen: 4.988 ug/l

RT: 12.40 min Scan# 1773

Delta R.T. 0.00 min

Lab File: VN056522.D

Acq: 27 Jun 2019 20:49

Instrument :

MSVOA_N

ClientSampleId :

0.05ppb LOD


Tot Ion: 95 Resp: 14227

Ion Ratio Lower Upper

95 100

174 84.9 0.0 167.8




176 78.5 0.0 158.0



Abundance Ion 95.00 (94.70 to 95.70): VN056516.D (-1763) (-)

Ion 174.00 (173.70 to 174.70): VN056516.D (-1763) (-)

Ion 176.00 (175.70 to 176.70): VN056516.D (-1763) (-)



#8

Chlorobenzene-d5

Concen: 5.000 ug/l

RT: 11.41 min Scan# 1628

Delta R.T. 0.00 min

Lab File: VN056522.D

Acq: 27 Jun 2019 20:49


Tgt Ion: 117 Resp: 55123

Ion Ratio Lower Upper

117 100

82 53.7 43.0 64.4



119 31.1 24.7 37.1




Abundance Ion 117.00 (116.70 to 117.70): VN056516.D (-1617) (-)

Ion 82.00 (81.70 to 82.70): VN056516.D (-1617) (-)

Ion 119.00 (118.70 to 119.70): VN056516.D (-1617) (-)





#9

1,4-Dichlorobenzene-d4

Concen: 5.000 ug/l

RT: 13.35 min Scan# 1864

Delta R.T. 0.00 min

Lab File: VN056522.D


Acq: 27 Jun 2019 20:49

Instrument :

MSVOA_N

ClientSampleId :

0.05ppb LOD



Tot Ion:152 Resp: 8935

Ion Ratio Lower Upper

152 100

115 54.0 27.0 80.8

150 155.0 0.0 306.6

Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V

