

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062719\
 Data File : VN056512.D
 Acq On : 27 Jun 2019 14:12
 Operator : JC/SP
 Sample : VSSTDICC0.1
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC0.1

Quant Time: Jun 28 03:20:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:15:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	56804	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	95109	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	55547	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	8214	5.00	ug/l	0.00
System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.01	65	964	0.14	ug/l	0.00
Spiked Amount 5.000	Range 81 - 118		Recovery	=	2.80%#	
5) Dibromofluoromethane	7.59	113	752	0.15	ug/l	0.00
Spiked Amount 5.000	Range 80 - 119		Recovery	=	3.00%#	
6) Toluene-d8	10.09	98	1510	0.13	ug/l	0.00
Spiked Amount 5.000	Range 89 - 112		Recovery	=	2.60%#	
7) 4-Bromofluorobenzene	12.40	95	482	0.18	ug/l	0.00
Spiked Amount 5.000	Range 85 - 114		Recovery	=	3.60%#	
Target Compounds						
2) Vinyl Chloride	2.17	62	646	0.115	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062719\
Data File : VN056512.D
Acq On : 27 Jun 2019 14:12
Operator : JC/SP
Sample : VSSTDICC0.1
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC0.1

Quant Time: Jun 28 03:20:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 03:15:11 2019
Response via : Initial Calibration

