

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062719\
 Data File : VN056514.D
 Acq On : 27 Jun 2019 14:59
 Operator : JC/SP
 Sample : VSSTDICC001
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC001

Quant Time: Jun 28 03:21:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:15:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	57163	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	97879	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	58186	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	8999	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	7156	1.05	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	21.00%#
5) Dibromofluoromethane	7.59	113	5358	1.07	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	21.40%#
6) Toluene-d8	10.09	98	11977	1.04	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	20.80%#
7) 4-Bromofluorobenzene	12.40	95	3050	1.09	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	21.80%#

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.16	62	5406	0.952	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062719\
Data File : VN056514.D
Acq On : 27 Jun 2019 14:59
Operator : JC/SP
Sample : VSSTDICC001
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC001

Quant Time: Jun 28 03:21:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 03:15:11 2019
Response via : Initial Calibration

