

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062719\
 Data File : VN056522.D
 Acq On : 27 Jun 2019 20:49
 Operator : JC/SP
 Sample : 0.05ppb LOD
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 0.05ppb LOD

Quant Time: Jun 28 03:30:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	52041	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	89398	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	55123	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	8935	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.01	65	30846	4.84	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	96.80%
5) Dibromofluoromethane	7.59	113	22866	4.64	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	92.80%
6) Toluene-d8	10.09	98	52717	4.73	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	94.60%
7) 4-Bromofluorobenzene	12.40	95	14227	4.99	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	99.80%

Target Compounds					Qvalue
2) Vinyl Chloride	2.17	62	251	0.048 ug/l #	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062719\
Data File : VN056522.D
Acq On : 27 Jun 2019 20:49
Operator : JC/SP
Sample : 0.05ppb LOD
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
0.05ppb LOD

Quant Time: Jun 28 03:30:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 03:22:32 2019
Response via : Initial Calibration

