

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062719\
 Data File : VN056529.D
 Acq On : 27 Jun 2019 23:33
 Operator : JC/SP
 Sample : VN0627WBSD01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0627WBSD01

Quant Time: Jun 28 03:30:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	50615	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	88826	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	53895	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	8541	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.01	65	31774	5.12	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	102.40%
5) Dibromofluoromethane	7.59	113	23535	4.81	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	96.20%
6) Toluene-d8	10.09	98	53049	4.79	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	95.80%
7) 4-Bromofluorobenzene	12.39	95	14087	4.97	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	99.40%

Target Compounds					Qvalue
2) Vinyl Chloride	2.16	62	26401	5.218 ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062719\
Data File : VN056529.D
Acq On : 27 Jun 2019 23:33
Operator : JC/SP
Sample : VN0627WBSD01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD01

Quant Time: Jun 28 03:30:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 03:22:32 2019
Response via : Initial Calibration

