

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062719\
 Data File : VN056533.D
 Acq On : 28 Jun 2019 1:07
 Operator : JC/SP
 Sample : K3519-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-19-JMW0882

Quant Time: Jun 28 03:31:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	51336	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	90044	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	60255	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	14376	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.01	65	31109	4.95	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	99.00%
5) Dibromofluoromethane	7.59	113	23336	4.70	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	94.00%
6) Toluene-d8	10.09	98	55561	4.95	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	99.00%
7) 4-Bromofluorobenzene	12.39	95	19817	6.90	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	138.00%#

Target Compounds					Qvalue	
2) Vinyl Chloride	2.16	62	260	0.051	ug/l	# 51

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062719\
 Data File : VN056533.D
 Acq On : 28 Jun 2019 1:07
 Operator : JC/SP
 Sample : K3519-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampled :
 SP-19-JMW0882

Quant Time: Jun 28 03:31:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

