

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062719\
 Data File : VN056538.D
 Acq On : 28 Jun 2019 3:04
 Operator : JC/SP
 Sample : VSSTDCCC005
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDCCC005EC

Quant Time: Jun 28 03:47:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 03:22:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	50805	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	89320	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	54532	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	8886	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.01	65	31466	5.06	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	101.20%
5) Dibromofluoromethane	7.59	113	23377	4.75	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	95.00%
6) Toluene-d8	10.09	98	53316	4.79	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	95.80%
7) 4-Bromofluorobenzene	12.39	95	14200	4.98	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	99.60%

Target Compounds					Qvalue
2) Vinyl Chloride	2.16	62	26156	5.150 ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062719\
Data File : VN056538.D
Acq On : 28 Jun 2019 3:04
Operator : JC/SP
Sample : VSSTDCCC005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDCCC005EC

Quant Time: Jun 28 03:47:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062719SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 03:22:32 2019
Response via : Initial Calibration

