

Method Path : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\
 Method File : 82N062719SIM-VINYL.M
 Title : SW846 8260
 Last Update : Mon Jul 01 15:45:13 2019
 Response Via : Initial Calibration

Calibration Files

0.1 =VN056512.D 0.5 =VN056513.D 1 =VN056514.D
 5 =VN056516.D 8 =VN056517.D 2 =VN056515.D

	Compound	0.1	0.5	1	5	8	2	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) C	Vinyl Chloride	0.569	0.496	0.473	0.497	0.490	0.474	0.500	7.07#
3) S	1,2-Dichloroethan		0.643	0.626	0.596	0.584	0.612	0.612	3.81
4) I	1,4-Difluorobenzene	-----ISTD-----							
5) S	Dibromofluorometh		0.327	0.274	0.256	0.255	0.266	0.276	10.73
6) S	Toluene-d8		0.722	0.612	0.590	0.592	0.603	0.624	8.95
7) S	4-Bromofluorobenz		0.186	0.156	0.143	0.158	0.154	0.160	9.96
8) I	Chlorobenzene-d5	-----ISTD-----							
9) I	1,4-Dichlorobenzene-d	-----ISTD-----							

(#) = Out of Range