

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073186.D
 Acq On : 28 Jun 2022 02:12
 Operator : JC\MD
 Sample : VN0627WBSD02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0627WBSD02

Quant Time: Jun 28 06:44:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	273741	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	486515	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	428891	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	188800	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	198508	53.638	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.280%	
35) Dibromofluoromethane	7.951	113	152427	50.689	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.380%	
50) Toluene-d8	10.380	98	555632	49.657	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.320%	
62) 4-Bromofluorobenzene	12.669	95	206400	51.298	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.600%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.052	85	46099	17.687	ug/l	98
3) Chloromethane	2.269	50	46066	16.824	ug/l	95
4) Vinyl Chloride	2.410	62	43958	18.807	ug/l	99
5) Bromomethane	2.804	94	21842	23.738	ug/l	89
6) Chloroethane	2.969	64	28297	19.351	ug/l	95
7) Trichlorofluoromethane	3.322	101	85556	22.477	ug/l	96
8) Diethyl Ether	3.769	74	39152	19.930	ug/l	82
9) 1,1,2-Trichlorotrifluo...	4.146	101	49564	18.838	ug/l	98
10) Methyl Iodide	4.351	142	54093	17.937	ug/l	96
11) Tert butyl alcohol	5.287	59	98378	102.907	ug/l	99
12) 1,1-Dichloroethene	4.116	96	48105	18.355	ug/l	94
13) Acrolein	3.987	56	13368	35.505	ug/l	94
14) Allyl chloride	4.763	41	103486	18.481	ug/l #	94
15) Acrylonitrile	5.469	53	227915	103.421	ug/l	98
16) Acetone	4.222	43	212383	104.398	ug/l	97
17) Carbon Disulfide	4.463	76	88227	13.139	ug/l	99
18) Methyl Acetate	4.781	43	117676	21.722	ug/l	91
19) Methyl tert-butyl Ether	5.534	73	214199	21.762	ug/l	100
20) Methylene Chloride	5.022	84	63157	19.869	ug/l	87
21) trans-1,2-Dichloroethene	5.510	96	49945	17.519	ug/l	93
22) Diisopropyl ether	6.416	45	223861	21.445	ug/l	96
23) Vinyl Acetate	6.351	43	993161	103.067	ug/l	94
24) 1,1-Dichloroethane	6.304	63	114627	20.945	ug/l	97
25) 2-Butanone	7.263	43	337731	104.618	ug/l	91
26) 2,2-Dichloropropane	7.251	77	70473	15.109	ug/l	98
27) cis-1,2-Dichloroethene	7.251	96	68305	20.054	ug/l	99
28) Bromochloromethane	7.586	49	45076	19.916	ug/l	86
29) Tetrahydrofuran	7.610	42	223563	104.871	ug/l	88
30) Chloroform	7.745	83	116542	21.035	ug/l	100
31) Cyclohexane	8.022	56	92071	18.391	ug/l	94
32) 1,1,1-Trichloroethane	7.939	97	99902	20.548	ug/l	100
36) 1,1-Dichloropropene	8.151	75	76088	18.700	ug/l	96
37) Ethyl Acetate	7.339	43	134427	20.655	ug/l	95
38) Carbon Tetrachloride	8.134	117	81638	19.075	ug/l	99
39) Methylcyclohexane	9.392	83	81516	16.621	ug/l	88
40) Benzene	8.386	78	248387	19.424	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073186.D
 Acq On : 28 Jun 2022 02:12
 Operator : JC\MD
 Sample : VN0627WBSD02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0627WBSD02

Quant Time: Jun 28 06:44:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.569	41	64963	21.049	ug/l #	88
42) 1,2-Dichloroethane	8.463	62	93655	20.620	ug/l	99
43) Isopropyl Acetate	8.492	43	195115	20.972	ug/l	94
44) Trichloroethene	9.145	130	60673	18.723	ug/l	97
45) 1,2-Dichloropropane	9.428	63	66925	20.475	ug/l	100
46) Dibromomethane	9.510	93	44575	20.168	ug/l	94
47) Bromodichloromethane	9.698	83	89732	20.429	ug/l	97
48) Methyl methacrylate	9.498	41	93101	21.918	ug/l	88
49) 1,4-Dioxane	9.504	88	35988	376.550	ug/l #	90
51) 4-Methyl-2-Pentanone	10.269	43	654177	107.829	ug/l	92
52) Toluene	10.439	92	155722	19.835	ug/l	99
53) t-1,3-Dichloropropene	10.657	75	92395	19.559	ug/l	99
54) cis-1,3-Dichloropropene	10.128	75	98516	19.093	ug/l	93
55) 1,1,2-Trichloroethane	10.833	97	68940	19.999	ug/l	99
56) Ethyl methacrylate	10.698	69	115110	20.833	ug/l #	87
57) 1,3-Dichloropropane	10.980	76	118362	20.696	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.980	63	261716	107.288	ug/l	96
59) 2-Hexanone	11.022	43	501903	106.963	ug/l	90
60) Dibromochloromethane	11.175	129	66281	20.345	ug/l	99
61) 1,2-Dibromoethane	11.280	107	69137	20.344	ug/l	98
64) Tetrachloroethene	10.916	164	59873	22.947	ug/l	96
65) Chlorobenzene	11.704	112	168362	20.100	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.780	131	65853	20.657	ug/l	98
67) Ethyl Benzene	11.780	91	299438	20.308	ug/l	100
68) m/p-Xylenes	11.892	106	222883	38.540	ug/l	95
69) o-Xylene	12.216	106	119225	20.632	ug/l	100
70) Styrene	12.233	104	188107	20.270	ug/l	99
71) Bromoform	12.392	173	47810	19.890	ug/l #	96
73) Isopropylbenzene	12.516	105	307550	21.056	ug/l	99
74) N-amyl acetate	12.327	43	156825	20.690	ug/l	90
75) 1,1,2,2-Tetrachloroethane	12.763	83	106626	20.270	ug/l	100
76) 1,2,3-Trichloropropane	12.821	75	137553	30.050	ug/l #	1
77) Bromobenzene	12.798	156	70302	20.091	ug/l	87
78) n-propylbenzene	12.857	91	333053	20.730	ug/l	98
79) 2-Chlorotoluene	12.945	91	211099	20.291	ug/l	99
80) 1,3,5-Trimethylbenzene	12.998	105	250357	20.867	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.563	75	29936	17.662	ug/l	90
82) 4-Chlorotoluene	13.039	91	200089	20.328	ug/l	99
83) tert-Butylbenzene	13.257	119	230841	21.008	ug/l	97
84) 1,2,4-Trimethylbenzene	13.304	105	248782	20.819	ug/l	99
85) sec-Butylbenzene	13.439	105	304572	20.779	ug/l	99
86) p-Isopropyltoluene	13.551	119	246982	20.812	ug/l	100
87) 1,3-Dichlorobenzene	13.551	146	126274	20.213	ug/l	99
88) 1,4-Dichlorobenzene	13.633	146	123233	19.490	ug/l	96
89) n-Butylbenzene	13.880	91	196261	20.254	ug/l	99
90) Hexachloroethane	14.145	117	47098	21.052	ug/l	86
91) 1,2-Dichlorobenzene	13.921	146	132855	20.393	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.539	75	28144	21.169	ug/l	90
93) 1,2,4-Trichlorobenzene	15.192	180	69382	19.243	ug/l	96
94) Hexachlorobutadiene	15.298	225	29697	18.928	ug/l	97
95) Naphthalene	15.433	128	294571	20.197	ug/l	99
96) 1,2,3-Trichlorobenzene	15.621	180	73165	19.004	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
Data File : VN073186.D
Acq On : 28 Jun 2022 02:12
Operator : JC\MD
Sample : VN0627WBSD02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02

Quant Time: Jun 28 06:44:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.016 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion:168 Resp: 273741
Ion Ratio Lower Upper
168 100
99 61.2 42.3 63.5



#2
Dichlorodifluoromethane
Concen: 17.687 ug/l
RT: 2.052 min Scan# 28
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 85 Resp: 46099
Ion Ratio Lower Upper
85 100
87 32.7 15.7 47.1









#3
Chloromethane
Concen: 16.824 ug/l
RT: 2.269 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 50 Resp: 46066
Ion Ratio Lower Upper
50 100
52 34.9 25.7 38.5



#4
Vinyl Chloride
Concen: 18.807 ug/l
RT: 2.410 min Scan# 89
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 62 Resp: 43958
Ion Ratio Lower Upper
62 100
64 30.5 25.0 37.4









#5
Bromomethane
Concen: 23.738 ug/l
RT: 2.804 min Scan# 11
Delta R.T. 0.017 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02

Tgt Ion: 94 Resp: 21842
Ion Ratio Lower Upper
94 100
96 84.7 76.0 114.0



#6
Chloroethane
Concen: 19.351 ug/l
RT: 2.969 min Scan# 184
Delta R.T. 0.018 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 64 Resp: 28297
Ion Ratio Lower Upper
64 100
66 30.1 26.2 39.4









#7
Trichlorofluoromethane
Concen: 22.477 ug/l
RT: 3.322 min Scan# 241
Delta R.T. 0.018 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion:101 Resp: 85556
Ion Ratio Lower Upper
101 100
103 67.3 51.5 77.3



#8
Diethyl Ether
Concen: 19.930 ug/l
RT: 3.769 min Scan# 320
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 74 Resp: 39152
Ion Ratio Lower Upper
74 100
45 102.5 42.9 128.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.838 ug/l
RT: 4.146 min Scan# 382
Delta R.T. 0.012 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02

Tgt Ion:101 Resp: 49564
Ion Ratio Lower Upper
101 100
85 45.2 34.7 52.1
151 80.4 63.0 94.4



#10
Methyl Iodide
Concen: 17.937 ug/l
RT: 4.351 min Scan# 419
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion:142 Resp: 54093
Ion Ratio Lower Upper
142 100
127 42.4 31.3 46.9
141 13.8 11.2 16.8









#11
Tert butyl alcohol
Concen: 102.907 ug/l
RT: 5.287 min Scan# 51
Delta R.T. -0.018 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02

Tgt Ion: 59 Resp: 98378
Ion Ratio Lower Upper
59 100
57 10.6 8.3 12.5



#12
1,1-Dichloroethene
Concen: 18.355 ug/l
RT: 4.116 min Scan# 379
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 96 Resp: 48105
Ion Ratio Lower Upper
96 100
61 151.9 127.6 191.4
98 60.0 51.2 76.8









#13
Acrolein
Concen: 35.505 ug/l
RT: 3.987 min Scan# 31
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02

Tgt Ion: 56 Resp: 13368
Ion Ratio Lower Upper
56 100
55 73.0 54.8 82.2



#14
Allyl chloride
Concen: 18.481 ug/l
RT: 4.763 min Scan# 489
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 41 Resp: 103486
Ion Ratio Lower Upper
41 100
39 64.5 53.2 79.8
76 29.4 29.9 44.9#









#15
Acrylonitrile
Concen: 103.421 ug/l
RT: 5.469 min Scan# 609
Delta R.T. -0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02

Tgt Ion: 53 Resp: 227915




Ion	Ratio	Lower	Upper
53	100		
52	84.6	66.4	99.6
51	36.5	28.5	42.7




#16
Acetone
Concen: 104.398 ug/l
RT: 4.222 min Scan# 397
Delta R.T. -0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 43 Resp: 212383

Ion	Ratio	Lower	Upper
43	100		
58	30.7	26.0	39.0





#17

Carbon Disulfide

Concen: 13.139 ug/l

RT: 4.463 min Scan# 41

Delta R.T. 0.006 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :





VN0627WBSD02

Tgt Ion: 76 Resp: 88227

Ion Ratio Lower Upper

76 100

78 8.8 7.2 10.8



#18

Methyl Acetate

Concen: 21.722 ug/l

RT: 4.781 min Scan# 492

Delta R.T. 0.000 min

Lab File: VN073186.D




Acq: 28 Jun 2022 02:12


Tgt Ion: 43 Resp: 117676

Ion Ratio Lower Upper

43 100

74 22.1 21.6 32.4









#19
Methyl tert-butyl Ether
Concen: 21.762 ug/l
RT: 5.534 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 73 Resp: 214199
Ion Ratio Lower Upper
73 100
57 22.0 17.6 26.4



#20
Methylene Chloride
Concen: 19.869 ug/l
RT: 5.022 min Scan# 533
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 84 Resp: 63157
Ion Ratio Lower Upper
84 100
49 132.3 88.3 132.5
51 34.8 28.2 42.2
86 59.6 50.6 76.0









#21
trans-1,2-Dichloroethene
Concen: 17.519 ug/l
RT: 5.510 min Scan# 616
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 96 Resp: 49945
Ion Ratio Lower Upper
96 100
61 139.3 105.8 158.8
98 55.8 50.3 75.5



#22
Diisopropyl ether
Concen: 21.445 ug/l
RT: 6.416 min Scan# 770
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 45 Resp: 223861
Ion Ratio Lower Upper
45 100
43 61.4 46.6 70.0
87 27.2 24.3 36.5
59 10.9 9.6 14.4









#23
 Vinyl Acetate
 Concen: 103.067 ug/l
 RT: 6.351 min Scan# 759
 Delta R.T. 0.000 min
 Lab File: VN073186.D
 Acq: 28 Jun 2022 02:12




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0627WBSD02


Tgt Ion: 43 Resp: 993161
 Ion Ratio Lower Upper
 43 100
 86 10.3 10.0 15.0



#24
 1,1-Dichloroethane
 Concen: 20.945 ug/l
 RT: 6.304 min Scan# 751
 Delta R.T. -0.006 min
 Lab File: VN073186.D
 Acq: 28 Jun 2022 02:12

Tgt Ion: 63 Resp: 114627
 Ion Ratio Lower Upper
 63 100
 98 8.3 3.5 10.4
 100 5.0 2.3 6.8





#25

2-Butanone

Concen: 104.618 ug/l

RT: 7.263 min Scan# 914

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :





VN0627WBSD02

Tgt Ion: 43 Resp: 337731

Ion Ratio Lower Upper

43 100

72 24.8 23.7 35.5



#26

2,2-Dichloropropane

Concen: 15.109 ug/l

RT: 7.251 min Scan# 912

Delta R.T. 0.000 min

Lab File: VN073186.D




Acq: 28 Jun 2022 02:12


Tgt Ion: 77 Resp: 70473

Ion Ratio Lower Upper

77 100

97 21.9 11.4 34.2





#27

cis-1,2-Dichloroethene

Concen: 20.054 ug/l

RT: 7.251 min Scan# 911

Delta R.T. 0.006 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :

VN0627WBSD02





Tgt Ion: 96 Resp: 68305

Ion Ratio Lower Upper

96 100

61 141.6 0.0 280.6

98 64.3 0.0 128.4



#28

Bromochloromethane

Concen: 19.916 ug/l

RT: 7.586 min Scan# 969

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12




Tgt Ion: 49 Resp: 45076

Ion Ratio Lower Upper

49 100

129 1.9 0.0 4.6

130 74.3 69.7 104.5









#29
Tetrahydrofuran
Concen: 104.871 ug/l
RT: 7.610 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 42 Resp: 223563
Ion Ratio Lower Upper
42 100
72 41.1 39.7 59.5
71 38.1 36.6 55.0



#30
Chloroform
Concen: 21.035 ug/l
RT: 7.745 min Scan# 996
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 83 Resp: 116542
Ion Ratio Lower Upper
83 100
85 64.7 51.8 77.8









#31
Cyclohexane
Concen: 18.391 ug/l
RT: 8.022 min Scan# 1043
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 56 Resp: 92071
Ion Ratio Lower Upper
56 100
69 36.5 27.8 41.8
84 89.0 76.9 115.3



#32
1,1,1-Trichloroethane
Concen: 20.548 ug/l
RT: 7.939 min Scan# 1029
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 97 Resp: 99902
Ion Ratio Lower Upper
97 100
99 63.0 50.6 76.0
61 45.2 36.0 54.0





#33

1,2-Dichloroethane-d4

Concen: 53.638 ug/l

RT: 8.369 min Scan# 1102

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :





VN0627WBSD02

Tgt Ion: 65 Resp: 198508

Ion Ratio Lower Upper

65 100

67 50.5 0.0 104.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.904 min Scan# 1193

Delta R.T. 0.006 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12




Tgt Ion: 114 Resp: 486515


Ion Ratio Lower Upper

114 100

63 20.8 0.0 40.0

88 15.3 0.0 32.6









#35
Dibromofluoromethane
Concen: 50.689 ug/l
RT: 7.951 min Scan# 1031
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02

Tgt Ion:113 Resp: 152427
Ion Ratio Lower Upper
113 100
111 102.8 81.4 122.0
192 19.8 17.0 25.6



#36
1,1-Dichloropropene
Concen: 18.700 ug/l
RT: 8.151 min Scan# 1065
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 75 Resp: 76088
Ion Ratio Lower Upper
75 100
110 33.7 18.3 54.8
77 30.1 25.1 37.7









#37
Ethyl Acetate
Concen: 20.655 ug/l
RT: 7.339 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 43 Resp: 134427
Ion Ratio Lower Upper
43 100
61 12.7 11.8 17.6
70 9.5 9.3 13.9



#38
Carbon Tetrachloride
Concen: 19.075 ug/l
RT: 8.134 min Scan# 1062
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 117 Resp: 81638
Ion Ratio Lower Upper
117 100
119 95.9 75.8 113.8
121 30.7 24.6 37.0





#39

Methylcyclohexane

Concen: 16.621 ug/l

RT: 9.392 min Scan# 1105

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :

VN0627WBSD02





Tgt Ion: 83 Resp: 81516

Ion Ratio Lower Upper

83 100

55 86.0 57.8 86.8

98 53.3 39.4 59.2



#40

Benzene

Concen: 19.424 ug/l

RT: 8.386 min Scan# 1105

Delta R.T. -0.006 min

Lab File: VN073186.D




Acq: 28 Jun 2022 02:12


Tgt Ion: 78 Resp: 248387

Ion Ratio Lower Upper

78 100

77 22.8 19.4 29.0









#41
Methacrylonitrile
Concen: 21.049 ug/l
RT: 7.569 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 41 Resp: 64963
Ion Ratio Lower Upper
41 100
39 61.7 48.6 73.0
67 71.4 73.4 110.2#
52 30.3 27.7 41.5



#42
1,2-Dichloroethane
Concen: 20.620 ug/l
RT: 8.463 min Scan# 1118
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 62 Resp: 93655
Ion Ratio Lower Upper
62 100
98 9.2 0.0 19.4









#43
Isopropyl Acetate
Concen: 20.972 ug/l
RT: 8.492 min Scan# 1123
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 43 Resp: 195115
Ion Ratio Lower Upper
43 100
61 24.3 21.9 32.9
87 13.0 12.0 18.0



#44
Trichloroethene
Concen: 18.723 ug/l
RT: 9.145 min Scan# 1234
Delta R.T. -0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion:130 Resp: 60673
Ion Ratio Lower Upper
130 100
95 96.4 0.0 186.8





#45

1,2-Dichloropropane

Concen: 20.475 ug/l

RT: 9.428 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :





VN0627WBSD02

Tgt Ion: 63 Resp: 66925

Ion Ratio Lower Upper

63 100

65 31.1 24.9 37.3



#46

Dibromomethane

Concen: 20.168 ug/l

RT: 9.510 min Scan# 1296

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12




Tgt Ion: 93 Resp: 44575


Ion Ratio Lower Upper

93 100

95 82.9 66.9 100.3

174 98.4 86.6 130.0









#47
Bromodichloromethane
Concen: 20.429 ug/l
RT: 9.698 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 83 Resp: 89732
Ion Ratio Lower Upper
83 100
85 65.4 50.5 75.7
127 9.3 6.8 10.2



#48
Methyl methacrylate
Concen: 21.918 ug/l
RT: 9.498 min Scan# 1294
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 41 Resp: 93101
Ion Ratio Lower Upper
41 100
69 76.5 74.7 112.1
39 58.5 45.6 68.4









#49
1,4-Dioxane
Concen: 376.550 ug/l
RT: 9.504 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02

Tgt Ion: 88 Resp: 35988
Ion Ratio Lower Upper
88 100
43 32.5 20.2 30.4#
58 74.3 54.2 81.4



#50
Toluene-d8
Concen: 49.657 ug/l
RT: 10.380 min Scan# 1444
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 98 Resp: 555632
Ion Ratio Lower Upper
98 100
100 65.5 52.6 78.8









#51
4-Methyl-2-Pentanone
Concen: 107.829 ug/l
RT: 10.269 min Scan# 1425
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02

Tgt Ion: 43 Resp: 654177
Ion Ratio Lower Upper
43 100
58 37.3 34.1 51.1



#52
Toluene
Concen: 19.835 ug/l
RT: 10.439 min Scan# 1454
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 92 Resp: 155722
Ion Ratio Lower Upper
92 100
91 171.5 138.6 207.8









#53
t-1,3-Dichloropropene
Concen: 19.559 ug/l
RT: 10.657 min Scan# 1491
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 75 Resp: 92395
Ion Ratio Lower Upper
75 100
77 32.6 25.6 38.4



#54
cis-1,3-Dichloropropene
Concen: 19.093 ug/l
RT: 10.128 min Scan# 1401
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 75 Resp: 98516
Ion Ratio Lower Upper
75 100
77 31.3 25.2 37.8
39 52.0 35.7 53.5









#55
1,1,2-Trichloroethane
Concen: 19.999 ug/l
RT: 10.833 min Scan# 1521
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion:	97	Resp:	68940
Ion	Ratio	Lower	Upper
97	100		
83	84.1	66.2	99.4
85	54.1	43.9	65.9
99	64.6	50.6	76.0



#56
Ethyl methacrylate
Concen: 20.833 ug/l
RT: 10.698 min Scan# 1498
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion:	69	Resp:	115110
Ion	Ratio	Lower	Upper
69	100		
41	71.8	47.8	71.6#
39	39.0	27.3	40.9





#57

1,3-Dichloropropane

Concen: 20.696 ug/l

RT: 10.980 min Scan# 1546

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :





VN0627WBSD02

Tgt Ion: 76 Resp: 118362

Ion Ratio Lower Upper

76 100

78 31.8 26.2 39.2



#58

2-Chloroethyl Vinyl ether

Concen: 107.288 ug/l

RT: 9.980 min Scan# 1376

Delta R.T. 0.000 min

Lab File: VN073186.D




Acq: 28 Jun 2022 02:12


Tgt Ion: 63 Resp: 261716

Ion Ratio Lower Upper

63 100

106 28.7 24.6 36.8









#59
2-Hexanone
Concen: 106.963 ug/l
RT: 11.022 min Scan# 1553
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 43 Resp: 501903
Ion Ratio Lower Upper
43 100
58 51.6 29.7 89.1



#60
Dibromochloromethane
Concen: 20.345 ug/l
RT: 11.175 min Scan# 1579
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 129 Resp: 66281
Ion Ratio Lower Upper
129 100
127 76.8 38.8 116.4









#61
1,2-Dibromoethane
Concen: 20.344 ug/l
RT: 11.280 min Scan# 1597
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion:107 Resp: 69137
Ion Ratio Lower Upper
107 100
109 95.6 75.0 112.6



#62
4-Bromofluorobenzene
Concen: 51.298 ug/l
RT: 12.669 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 95 Resp: 206400
Ion Ratio Lower Upper
95 100
174 75.1 0.0 167.6
176 72.2 0.0 163.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.680 min Scan# 1166
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion:117 Resp: 428891
Ion Ratio Lower Upper
117 100
82 55.2 44.2 66.2
119 32.7 25.4 38.0



#64
Tetrachloroethene
Concen: 22.947 ug/l
RT: 10.916 min Scan# 1535
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion:164 Resp: 59873
Ion Ratio Lower Upper
164 100
166 122.2 104.2 156.2
129 89.0 73.9 110.9
131 85.6 69.4 104.2









#65
Chlorobenzene
Concen: 20.100 ug/l
RT: 11.704 min Scan# 1669
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion:112 Resp: 168362
Ion Ratio Lower Upper
112 100
114 34.3 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 20.657 ug/l
RT: 11.780 min Scan# 1682
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion:131 Resp: 65853
Ion Ratio Lower Upper
131 100
133 96.1 47.8 143.4
119 64.3 33.4 100.2









#67
Ethyl Benzene
Concen: 20.308 ug/l
RT: 11.780 min Scan# 1
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 91 Resp: 299438
Ion Ratio Lower Upper
91 100
106 31.3 24.9 37.3



#68
m/p-Xylenes
Concen: 38.540 ug/l
RT: 11.892 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 106 Resp: 222883
Ion Ratio Lower Upper
106 100
91 206.2 158.9 238.3









#69
o-Xylene
Concen: 20.632 ug/l
RT: 12.216 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion:106 Resp: 119225
Ion Ratio Lower Upper
106 100
91 212.4 106.0 318.0



#70
Styrene
Concen: 20.270 ug/l
RT: 12.233 min Scan# 1759
Delta R.T. 0.006 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion:104 Resp: 188107
Ion Ratio Lower Upper
104 100
78 51.1 39.5 59.3
103 55.3 44.0 66.0





#71

Bromoform

Concen: 19.890 ug/l

RT: 12.392 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :

VN0627WBSD02





Tgt Ion:173 Resp: 47810

Ion Ratio Lower Upper

173 100

175 52.0 24.6 73.6

254 0.0 0.1 0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.610 min Scan# 1993

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12




Tgt Ion:152 Resp: 188800


Ion Ratio Lower Upper

152 100

115 62.3 28.9 86.7

150 165.9 0.0 356.2









#73
Isopropylbenzene
Concen: 21.056 ug/l
RT: 12.516 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion:105 Resp: 307550
Ion Ratio Lower Upper
105 100
120 26.5 13.5 40.4



#74
N-ethyl acetate
Concen: 20.690 ug/l
RT: 12.327 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 43 Resp: 156825
Ion Ratio Lower Upper
43 100
70 39.6 39.0 58.4
55 22.8 21.1 31.7
61 23.4 21.1 31.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 20.270 ug/l
 RT: 12.763 min Scan# 1849
 Delta R.T. 0.000 min
 Lab File: VN073186.D
 Acq: 28 Jun 2022 02:12




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0627WBSD02

Tgt Ion: 83 Resp: 106626
 Ion Ratio Lower Upper
 83 100
 131 11.2 5.3 15.9
 85 64.7 32.3 96.9



#76
 1,2,3-Trichloropropane
 Concen: 30.050 ug/l
 RT: 12.821 min Scan# 1859
 Delta R.T. 0.006 min
 Lab File: VN073186.D
 Acq: 28 Jun 2022 02:12

Tgt Ion: 75 Resp: 137553
 Ion Ratio Lower Upper
 75 100
 77 0.0 98.5 295.5#









#77
Bromobenzene
Concen: 20.091 ug/l
RT: 12.798 min Scan# 1855
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion:156 Resp: 70302
Ion Ratio Lower Upper
156 100
77 229.1 100.5 301.5
158 100.9 48.9 146.6



#78
n-propylbenzene
Concen: 20.730 ug/l
RT: 12.857 min Scan# 1865
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 91 Resp: 333053
Ion Ratio Lower Upper
91 100
120 22.7 11.9 35.5





#79

2-Chlorotoluene

Concen: 20.291 ug/l

RT: 12.945 min Scan# 1880

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :





VN0627WBSD02

Tgt Ion: 91 Resp: 211099

Ion Ratio Lower Upper

91 100

126 33.9 17.3 51.9



#80

1,3,5-Trimethylbenzene

Concen: 20.867 ug/l

RT: 12.998 min Scan# 1889

Delta R.T. 0.000 min

Lab File: VN073186.D




Acq: 28 Jun 2022 02:12


Tgt Ion: 105 Resp: 250357

Ion Ratio Lower Upper

105 100

120 46.9 24.9 74.7









#81
trans-1,4-Dichloro-2-butene
Concen: 17.662 ug/l
RT: 12.563 min Scan# 1815
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion: 75 Resp: 29936
Ion Ratio Lower Upper
75 100
53 110.0 78.0 117.0
89 48.6 37.2 55.8



#82
4-Chlorotoluene
Concen: 20.328 ug/l
RT: 13.039 min Scan# 1896
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion: 91 Resp: 200089
Ion Ratio Lower Upper
91 100
126 33.8 17.2 51.6









#83
tert-Butylbenzene
Concen: 21.008 ug/l
RT: 13.257 min Scan# 1933
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12




Instrument :
MSVOA_N
ClientSampleId :
VN0627WBSD02


Tgt Ion:119 Resp: 230841
Ion Ratio Lower Upper
119 100
91 65.4 31.6 94.8
134 26.3 13.7 41.1



#84
1,2,4-Trimethylbenzene
Concen: 20.819 ug/l
RT: 13.304 min Scan# 1941
Delta R.T. 0.000 min
Lab File: VN073186.D
Acq: 28 Jun 2022 02:12

Tgt Ion:105 Resp: 248782
Ion Ratio Lower Upper
105 100
120 45.8 23.3 69.8









#85
 sec-Butylbenzene
 Concen: 20.779 ug/l
 RT: 13.439 min Scan# 1964
 Delta R.T. 0.000 min
 Lab File: VN073186.D
 Acq: 28 Jun 2022 02:12




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0627WBSD02


Tgt Ion:105 Resp: 304572
 Ion Ratio Lower Upper
 105 100
 134 20.4 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 20.812 ug/l
 RT: 13.551 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN073186.D
 Acq: 28 Jun 2022 02:12

Tgt Ion:119 Resp: 246982
 Ion Ratio Lower Upper
 119 100
 134 26.5 13.3 39.8
 91 23.0 11.3 33.9





#87

1,3-Dichlorobenzene

Concen: 20.213 ug/l

RT: 13.551 min Scan# 1983

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :

VN0627WBSD02





Tgt Ion:146 Resp: 126274

Ion Ratio Lower Upper

146 100

111 40.1 20.0 60.0

148 63.0 32.0 96.0



#88

1,4-Dichlorobenzene

Concen: 19.490 ug/l

RT: 13.633 min Scan# 1997

Delta R.T. 0.006 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12




Tgt Ion:146 Resp: 123233


Ion Ratio Lower Upper

146 100

111 41.5 19.1 57.1

148 64.8 31.4 94.3





#89

n-Butylbenzene

Concen: 20.254 ug/l

RT: 13.880 min Scan# 2038

Delta R.T. 0.006 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :

VN0627WBSD02





Tgt Ion: 91 Resp: 196261

Ion Ratio Lower Upper

91 100

92 53.2 26.9 80.7

134 26.2 13.6 40.8



#90

Hexachloroethane

Concen: 21.052 ug/l

RT: 14.145 min Scan# 2084

Delta R.T. 0.000 min

Lab File: VN073186.D




Acq: 28 Jun 2022 02:12


Tgt Ion: 117 Resp: 47098

Ion Ratio Lower Upper

117 100

201 68.7 40.6 122.0









#91
 1,2-Dichlorobenzene
 Concen: 20.393 ug/l
 RT: 13.921 min Scan# 2046
 Delta R.T. 0.000 min
 Lab File: VN073186.D
 Acq: 28 Jun 2022 02:12




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0627WBSD02


Tgt Ion:146 Resp: 132855
 Ion Ratio Lower Upper
 146 100
 111 43.2 20.9 62.8
 148 63.2 32.1 96.3



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 21.169 ug/l
 RT: 14.539 min Scan# 2151
 Delta R.T. -0.000 min
 Lab File: VN073186.D
 Acq: 28 Jun 2022 02:12

Tgt Ion: 75 Resp: 28144
 Ion Ratio Lower Upper
 75 100
 155 82.4 44.9 134.5
 157 104.7 59.0 176.8





#93

1,2,4-Trichlorobenzene

Concen: 19.243 ug/l

RT: 15.192 min Scan# 2262

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :

VN0627WBSD02





Tgt Ion:180 Resp: 69382

Ion Ratio Lower Upper

180 100

182 93.5 47.5 142.5

145 34.5 14.5 43.5



#94

Hexachlorobutadiene

Concen: 18.928 ug/l

RT: 15.298 min Scan# 2280

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12




Tgt Ion:225 Resp: 29697


Ion Ratio Lower Upper

225 100

223 61.5 32.4 97.2

227 65.4 31.8 95.4





#95

Naphthalene

Concen: 20.197 ug/l

RT: 15.433 min Scan# 2303

Delta R.T. 0.006 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Instrument :

MSVOA_N

ClientSampleId :

VN0627WBSD02





Tgt Ion:128 Resp: 294571

Ion Ratio Lower Upper

128 100

127 13.1 10.3 15.5

129 11.6 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 19.004 ug/l

RT: 15.621 min Scan# 2335

Delta R.T. 0.000 min

Lab File: VN073186.D

Acq: 28 Jun 2022 02:12

Tgt Ion:180 Resp: 73165

Ion Ratio Lower Upper

180 100

182 95.7 47.8 143.3

145 35.6 15.5 46.5

