

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073154.D
 Acq On : 27 Jun 2022 11:58
 Operator : JC\MD
 Sample : PB145841TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB145841TB

Quant Time: Jun 28 06:36:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	332735	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	534343	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	495214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	245526	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	200867	44.652	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.300%
35) Dibromofluoromethane	7.951	113	167326	50.663	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.320%
50) Toluene-d8	10.380	98	568118	46.228	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.460%
62) 4-Bromofluorobenzene	12.668	95	246390	55.755	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.520%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	28938	11.703	ug/l	98
20) Methylene Chloride	5.016	84	33710	8.589	ug/l	90
43) Isopropyl Acetate	8.498	43	44757	4.380	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
Data File : VN073154.D
Acq On : 27 Jun 2022 11:58
Operator : JC\MD
Sample : PB145841TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB145841TB

Quant Time: Jun 28 06:36:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

