

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073159.D
 Acq On : 27 Jun 2022 14:00
 Operator : JC\MD
 Sample : N3494-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-EXC-1-4-GR

Quant Time: Jun 28 06:37:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	305522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	498533	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	459343	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	227555	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	188038	45.523	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.040%		
35) Dibromofluoromethane	7.957	113	98921	32.103	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	64.200%#		
50) Toluene-d8	10.380	98	521128	45.450	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.900%		
62) 4-Bromofluorobenzene	12.669	95	228157	55.338	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	110.680%		
Target Compounds					Qvalue	
16) Acetone	4.234	43	79763	35.129	ug/l	99
20) Methylene Chloride	5.016	84	51392	14.420	ug/l	89
25) 2-Butanone	7.275	43	19720	5.473	ug/l	91
65) Chlorobenzene	11.710	112	52730	5.878	ug/l	97
91) 1,2-Dichlorobenzene	13.927	146	13762	1.753	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
Data File : VN073159.D
Acq On : 27 Jun 2022 14:00
Operator : JC\MD
Sample : N3494-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-EXC-1-4-GR

Quant Time: Jun 28 06:37:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

