

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073160.D
 Acq On : 27 Jun 2022 14:25
 Operator : JC\MD
 Sample : N3494-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-EXC-1-5

Quant Time: Jun 28 06:37:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	301195	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	485699	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.681	117	443402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	217412	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	180869	44.417	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.840%
35) Dibromofluoromethane	7.951	113	151307	50.401	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.800%
50) Toluene-d8	10.381	98	512588	45.887	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.780%
62) 4-Bromofluorobenzene	12.669	95	221606	55.169	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.340%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	94189	42.079	ug/l	96
20) Methylene Chloride	5.010	84	31391	8.842	ug/l #	93
65) Chlorobenzene	11.710	112	111453	12.871	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
Data File : VN073160.D
Acq On : 27 Jun 2022 14:25
Operator : JC\MD
Sample : N3494-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-EXC-1-5

Quant Time: Jun 28 06:37:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

