

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073163.D
 Acq On : 27 Jun 2022 15:38
 Operator : JC\MD
 Sample : N3494-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-EXC-1-6-GR

Quant Time: Jun 28 06:37:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	310283	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	512794	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	469118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	226771	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	191287	45.599	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.200%
35) Dibromofluoromethane	7.951	113	66956	21.125	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	42.240%#
50) Toluene-d8	10.380	98	538465	45.656	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.320%
62) 4-Bromofluorobenzene	12.669	95	232080	54.724	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.440%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	137311	59.547	ug/l	93
20) Methylene Chloride	5.016	84	110292	30.743	ug/l	95
25) 2-Butanone	7.275	43	27638	7.553	ug/l	92
95) Naphthalene	15.433	128	176653	10.084	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
Data File : VN073163.D
Acq On : 27 Jun 2022 15:38
Operator : JC\MD
Sample : N3494-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-EXC-1-6-GR

Quant Time: Jun 28 06:37:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

