

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073166.D
 Acq On : 27 Jun 2022 16:52
 Operator : JC\MD
 Sample : N3422-09
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-13

Quant Time: Jun 27 17:48:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	326941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	525431	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	516043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	439096	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	192517	43.554	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	87.100%		
35) Dibromofluoromethane	7.945	113	157432	48.476	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.960%		
50) Toluene-d8	10.375	98	580288	48.019	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.040%		
62) 4-Bromofluorobenzene	12.669	95	285569	65.717	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	131.440%#		
Target Compounds						Qvalue
16) Acetone	4.251	43	47741	19.649	ug/l	91
18) Methyl Acetate	4.775	43	39157	6.052	ug/l #	64
25) 2-Butanone	7.281	43	60198	15.613	ug/l	94
31) Cyclohexane	8.004	56	39366	6.584	ug/l #	85
39) Methylcyclohexane	9.386	83	111690	21.086	ug/l	95
52) Toluene	10.439	92	25289	2.983	ug/l	91
67) Ethyl Benzene	11.786	91	51209	2.886	ug/l	92
68) m/p-Xylenes	11.886	106	75700	10.879	ug/l	96
69) o-Xylene	12.216	106	45336	6.520	ug/l	97
80) 1,3,5-Trimethylbenzene	12.998	105	183980	6.594	ug/l	78
84) 1,2,4-Trimethylbenzene	13.304	105	393374	14.154	ug/l	88
86) p-Isopropyltoluene	13.551	119	318920	11.555	ug/l	94
95) Naphthalene	15.433	128	371622	10.955	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
Data File : VN073166.D
Acq On : 27 Jun 2022 16:52
Operator : JC\MD
Sample : N3422-09
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-1A-13

Quant Time: Jun 27 17:48:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

