

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073172.D
 Acq On : 27 Jun 2022 19:18
 Operator : JC\MD
 Sample : N3494-05RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-EXC-3-GRRE

Quant Time: Jun 28 06:39:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	334608	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	536013	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	473158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	209017	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	195329	43.178	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.360%
35) Dibromofluoromethane	7.951	113	71228	21.499	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	43.000%#
50) Toluene-d8	10.374	98	554892	45.011	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.020%
62) 4-Bromofluorobenzene	12.668	95	228615	51.572	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.140%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	79348	31.909	ug/l	93
20) Methylene Chloride	5.010	84	75053	19.310	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
Data File : VN073172.D
Acq On : 27 Jun 2022 19:18
Operator : JC\MD
Sample : N3494-05RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-EXC-3-GRRE

Quant Time: Jun 28 06:39:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

