

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073176.D
 Acq On : 27 Jun 2022 22:09
 Operator : JC\MD
 Sample : VN0627MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0627MBL02

Quant Time: Jun 28 06:40:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	335894	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	540582	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	502423	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	254341	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	203134	44.731	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.460%
35) Dibromofluoromethane	7.957	113	172654	51.673	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.340%
50) Toluene-d8	10.380	98	567031	45.607	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.220%
62) 4-Bromofluorobenzene	12.669	95	259301	58.000	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
Data File : VN073176.D
Acq On : 27 Jun 2022 22:09
Operator : JC\MD
Sample : VN0627MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0627MBL02

Quant Time: Jun 28 06:40:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

