

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073177.D
 Acq On : 27 Jun 2022 22:33
 Operator : JC\MD
 Sample : VN0627WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0627WBL02

Quant Time: Jun 28 06:40:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	351003	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	558389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	518427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	258407	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	211113	44.487	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.980%
35) Dibromofluoromethane	7.951	113	176920	51.261	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.520%
50) Toluene-d8	10.374	98	585426	45.585	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.160%
62) 4-Bromofluorobenzene	12.668	95	265007	57.386	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
Data File : VN073177.D
Acq On : 27 Jun 2022 22:33
Operator : JC\MD
Sample : VN0627WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0627WBL02

Quant Time: Jun 28 06:40:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

