

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073188.D
 Acq On : 28 Jun 2022 03:01
 Operator : JC\MD
 Sample : N3494-27
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-EXC-1-2

Quant Time: Jun 28 06:44:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	348125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	567808	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	514258	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	252598	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	205451	43.652	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.300%
35) Dibromofluoromethane	7.957	113	71735	20.440	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	40.880%#
50) Toluene-d8	10.375	98	592339	45.358	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.720%
62) 4-Bromofluorobenzene	12.669	95	255687	54.449	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.900%
Target Compounds						
					Qvalue	
16) Acetone	4.222	43	121778	47.070	ug/l	91
20) Methylene Chloride	5.016	84	108416	26.905	ug/l	94
25) 2-Butanone	7.269	43	29437	7.170	ug/l #	88
40) Benzene	8.392	78	50572	3.389	ug/l	96
65) Chlorobenzene	11.710	112	454170	45.222	ug/l	99
91) 1,2-Dichlorobenzene	13.927	146	15439	1.771	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
Data File : VN073188.D
Acq On : 28 Jun 2022 03:01
Operator : JC\MD
Sample : N3494-27
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-EXC-1-2

Quant Time: Jun 28 06:44:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

