

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073149.D
 Acq On : 27 Jun 2022 08:41
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 27 17:46:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.476	0.443	6.9	88	0.00
3 P	Chloromethane	0.535	0.412	23.0	83	0.00
4 C	Vinyl Chloride	0.427	0.374	12.4#	85	0.00
5 T	Bromomethane	0.168	0.166	1.2	90	0.00
6 T	Chloroethane	0.267	0.236	11.6	88	0.00
7 T	Trichlorofluoromethane	0.695	0.756	-8.8	117	0.00
8 T	Diethyl Ether	0.359	0.326	9.2	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.481	0.470	2.3	98	0.00
10 T	Methyl Iodide	0.551	0.457	17.1	82	0.00
11 T	Tert butyl alcohol	0.175	0.142	18.9	79	-0.02
12 CM	1,1-Dichloroethene	0.479	0.425	11.3#	87	0.00
13 T	Acrolein	0.069	0.020	71.0#	28#	0.00
14 T	Allyl chloride	1.023	0.938	8.3	92	0.00
15 T	Acrylonitrile	0.403	0.369	8.4	90	0.00
16 T	Acetone	0.372	0.387	-4.0	107	-0.01
17 T	Carbon Disulfide	1.227	0.875	28.7#	72	0.00
18 T	Methyl Acetate	0.990	0.897	9.4	95	0.00
19 T	Methyl tert-butyl Ether	1.798	1.799	-0.1	97	0.00
20 T	Methylene Chloride	0.628	0.531	15.4	90	0.00
21 T	trans-1,2-Dichloroethene	0.521	0.448	14.0	87	0.00
22 T	Diisopropyl ether	1.907	1.908	-0.1	97	-0.01
23 T	Vinyl Acetate	1.760	1.721	2.2	94	0.00
24 P	1,1-Dichloroethane	1.000	0.962	3.8	95	-0.01
25 T	2-Butanone	0.590	0.553	6.3	92	0.00
26 T	2,2-Dichloropropane	0.898	0.868	3.3	100	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.595	4.3	96	0.00
28 T	Bromochloromethane	0.413	0.410	0.7	95	0.00
29 T	Tetrahydrofuran	0.389	0.350	10.0	89	0.00
30 C	Chloroform	1.012	0.996	1.6#	96	0.00
31 T	Cyclohexane	0.914	0.768	16.0	89	0.00
32 T	1,1,1-Trichloroethane	0.888	0.869	2.1	95	0.00
33 S	1,2-Dichloroethane-d4	0.676	0.672	0.6	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.309	0.316	-2.3	98	0.00
36 T	1,1-Dichloropropene	0.418	0.397	5.0	91	0.00
37 T	Ethyl Acetate	0.669	0.656	1.9	96	-0.01
38 T	Carbon Tetrachloride	0.440	0.432	1.8	94	0.00
39 T	Methylcyclohexane	0.504	0.476	5.6	89	0.00
40 TM	Benzene	1.314	1.258	4.3	91	0.00
41 T	Methacrylonitrile	0.317	0.312	1.6	97	0.00
42 TM	1,2-Dichloroethane	0.467	0.481	-3.0	97	0.00
43 T	Isopropyl Acetate	0.956	0.981	-2.6	96	0.00
44 TM	Trichloroethene	0.333	0.319	4.2	93	0.00
45 C	1,2-Dichloropropane	0.336	0.339	-0.9#	97	0.00
46 T	Dibromomethane	0.227	0.227	0.0	94	0.00
47 T	Bromodichloromethane	0.451	0.470	-4.2	97	0.00
48 T	Methyl methacrylate	0.437	0.427	2.3	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
 Data File : VN073149.D
 Acq On : 27 Jun 2022 08:41
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 27 17:46:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.008	20.0	80	0.00
50 S	Toluene-d8	1.150	1.181	-2.7	97	0.00
51 T	4-Methyl-2-Pentanone	0.623	0.626	-0.5	94	0.00
52 CM	Toluene	0.807	0.797	1.2#	93	0.00
53 T	t-1,3-Dichloropropene	0.485	0.524	-8.0	97	0.00
54 T	cis-1,3-Dichloropropene	0.530	0.547	-3.2	96	0.00
55 T	1,1,2-Trichloroethane	0.354	0.343	3.1	96	0.00
56 T	Ethyl methacrylate	0.568	0.585	-3.0	95	0.00
57 T	1,3-Dichloropropane	0.588	0.580	1.4	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.251	0.263	-4.8	92	0.00
59 T	2-Hexanone	0.482	0.489	-1.5	94	0.00
60 T	Dibromochloromethane	0.335	0.364	-8.7	96	0.00
61 T	1,2-Dibromoethane	0.349	0.349	0.0	93	0.00
62 S	4-Bromofluorobenzene	0.414	0.450	-8.7	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.304	0.320	-5.3	103	0.00
65 PM	Chlorobenzene	0.976	0.965	1.1	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.377	-1.3	96	0.00
67 C	Ethyl Benzene	1.719	1.785	-3.8#	97	0.00
68 T	m/p-Xylenes	0.674	0.671	0.4	95	0.00
69 T	o-Xylene	0.674	0.682	-1.2	97	0.00
70 T	Styrene	1.082	1.108	-2.4	95	0.00
71 P	Bromoform	0.280	0.301	-7.5	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.868	3.819	1.3	99	0.00
74 T	N-amyl acetate	2.007	1.886	6.0	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.393	1.264	9.3	93	0.00
76 T	1,2,3-Trichloropropane	1.212	1.050	13.4	97	0.00
77 T	Bromobenzene	0.927	0.867	6.5	93	0.00
78 T	n-propylbenzene	4.255	4.433	-4.2	101	0.00
79 T	2-Chlorotoluene	2.755	2.659	3.5	98	0.00
80 T	1,3,5-Trimethylbenzene	3.177	3.152	0.8	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.449	0.448	0.2	95	0.00
82 T	4-Chlorotoluene	2.607	2.616	-0.3	99	0.00
83 T	tert-Butylbenzene	2.910	2.855	1.9	97	0.00
84 T	1,2,4-Trimethylbenzene	3.165	3.171	-0.2	99	0.00
85 T	sec-Butylbenzene	3.882	3.983	-2.6	100	0.00
86 T	p-Isopropyltoluene	3.143	3.281	-4.4	99	0.00
87 T	1,3-Dichlorobenzene	1.654	1.609	2.7	99	0.00
88 T	1,4-Dichlorobenzene	1.674	1.611	3.8	98	0.00
89 T	n-Butylbenzene	2.566	2.842	-10.8	104	0.00
90 T	Hexachloroethane	0.592	0.599	-1.2	98	0.00
91 T	1,2-Dichlorobenzene	1.725	1.640	4.9	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.376	0.315	16.2	89	0.00
93 T	1,2,4-Trichlorobenzene	0.955	0.956	-0.1	99	0.00
94 T	Hexachlorobutadiene	0.416	0.418	-0.5	101	0.00
95 T	Naphthalene	3.863	3.469	10.2	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062722\
Data File : VN073149.D
Acq On : 27 Jun 2022 08:41
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 27 17:46:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.020	0.930	8.8	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6