

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062723\
 Data File : VN078413.D
 Acq On : 27 Jun 2023 12:23
 Operator : JC\MD
 Sample : VN0627WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0627WBL01

Quant Time: Jun 28 07:07:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	431321	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	787680	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	691374	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	196054	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	299608	52.098	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.200%
35) Dibromofluoromethane	8.177	113	278572	55.934	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.860%
50) Toluene-d8	10.570	98	910163	47.404	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.800%
62) 4-Bromofluorobenzene	12.853	95	273654	45.847	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062723\
Data File : VN078413.D
Acq On : 27 Jun 2023 12:23
Operator : JC\MD
Sample : VN0627WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0627WBL01

Quant Time: Jun 28 07:07:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

