

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062723\
 Data File : VN078417.D
 Acq On : 27 Jun 2023 15:29
 Operator : JC\MD
 Sample : 03390-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TW-3

Quant Time: Jun 28 07:10:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/28/2023
 Supervised By : Mahesh Dadoda 06/28/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	455113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	835504	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	735383	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	206362	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	305459	50.339	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.680%			
35) Dibromofluoromethane	8.177	113	294464	55.740	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 111.480%			
50) Toluene-d8	10.577	98	938540	46.084	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 92.160%			
62) 4-Bromofluorobenzene	12.853	95	292454	46.193	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 92.380%			
Target Compounds						
16) Acetone	4.448	43	5336m	3.055	ug/l	
27) cis-1,2-Dichloroethene	7.494	96	12861	2.011	ug/l	95
30) Chloroform	7.977	83	27427	2.706	ug/l	86
44) Trichloroethene	9.359	130	9959	1.667	ug/l	91
64) Tetrachloroethene	11.112	164	1159727	236.390	ug/l	99
87) 1,3-Dichlorobenzene	13.741	146	2363	0.308	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062723\
Data File : VN078417.D
Acq On : 27 Jun 2023 15:29
Operator : JC\MD
Sample : 03390-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TW-3

Quant Time: Jun 28 07:10:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/28/2023
Supervised By :Mahesh Dadoda 06/28/2023

