

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062723\
 Data File : VN078418.D
 Acq On : 27 Jun 2023 15:53
 Operator : JC\MD
 Sample : 03376-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TW-1

Quant Time: Jun 28 07:11:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	447279	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	814732	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	737908	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	237328	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	305703	51.262	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.520%	
35) Dibromofluoromethane	8.177	113	292476	56.776	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 113.560%	
50) Toluene-d8	10.570	98	945929	47.631	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.260%	
62) 4-Bromofluorobenzene	12.853	95	312367	50.596	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.200%	
Target Compounds						
						Qvalue
30) Chloroform	7.977	83	47779	4.796	ug/l	87
64) Tetrachloroethene	11.106	164	149742	30.418	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062723\
Data File : VN078418.D
Acq On : 27 Jun 2023 15:53
Operator : JC\MD
Sample : 03376-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TW-1

Quant Time: Jun 28 07:11:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

