

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062723\
 Data File : VN078422.D
 Acq On : 27 Jun 2023 17:29
 Operator : JC\MD
 Sample : 03390-01DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TW-3DL

Quant Time: Jun 28 07:12:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	433224	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	791844	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	727166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	217653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	304691	52.749	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.500%
35) Dibromofluoromethane	8.177	113	285719	57.067	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	114.140%
50) Toluene-d8	10.570	98	935188	48.452	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.900%
62) 4-Bromofluorobenzene	12.859	95	311235	51.870	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.740%
Target Compounds						
					Qvalue	
64) Tetrachloroethene	11.112	164	195748	40.351	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062723\
Data File : VN078422.D
Acq On : 27 Jun 2023 17:29
Operator : JC\MD
Sample : 03390-01DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TW-3DL

Quant Time: Jun 28 07:12:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

