

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062723\
 Data File : VN078426.D
 Acq On : 27 Jun 2023 19:05
 Operator : JC\MD
 Sample : 03417-03 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1622

Quant Time: Jun 28 07:14:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	386139	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	719475	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	593554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	157866	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	280084	54.402	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.800%	
35) Dibromofluoromethane	8.171	113	245305	53.923	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.840%	
50) Toluene-d8	10.576	98	851643	48.561	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.120%	
62) 4-Bromofluorobenzene	12.853	95	220049	40.362	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 80.720%#	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	24229	16.350	ug/l	95
20) Methylene Chloride	5.294	84	3166	0.647	ug/l #	82
52) Toluene	10.641	92	38453	3.126	ug/l	100
67) Ethyl Benzene	11.965	91	7427	0.339	ug/l	96
68) m/p-Xylenes	12.070	106	10193	1.219	ug/l	99
69) o-Xylene	12.406	106	3659	0.446	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062723\
Data File : VN078426.D
Acq On : 27 Jun 2023 19:05
Operator : JC\MD
Sample : 03417-03 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1622

Quant Time: Jun 28 07:14:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

