

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
 Data File : VN087208.D
 Acq On : 27 Jun 2025 10:12
 Operator : JC\MD
 Sample : VN0627MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0627MBL01

Quant Time: Jun 27 23:39:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 27 05:55:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	142070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	275807	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	244362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	115263	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	109422	53.520	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.040%
35) Dibromofluoromethane	8.171	113	85174	47.767	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.540%
50) Toluene-d8	10.565	98	344830	46.758	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.520%
62) 4-Bromofluorobenzene	12.847	95	118704	46.387	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
Data File : VN087208.D
Acq On : 27 Jun 2025 10:12
Operator : JC\MD
Sample : VN0627MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0627MBL01

Quant Time: Jun 27 23:39:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 27 05:55:21 2025
Response via : Initial Calibration

