

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
 Data File : VN087216.D
 Acq On : 27 Jun 2025 13:10
 Operator : JC\MD
 Sample : Q2414-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Jun 27 23:43:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 27 05:55:21 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	125710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	293036	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	291791	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	134948	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	122877	67.922	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 135.840%#	
35) Dibromofluoromethane	8.177	113	78957	41.677	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 83.360%	
50) Toluene-d8	10.565	98	372856	47.585	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.180%	
62) 4-Bromofluorobenzene	12.847	95	143998	52.963	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.920%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	41565	36.633	ug/l	96
18) Methyl Acetate	5.053	43	24409	10.392	ug/l	99
20) Methylene Chloride	5.300	84	9848	4.482	ug/l	92
43) Isopropyl Acetate	8.694	43	10192	1.998	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
Data File : VN087216.D
Acq On : 27 Jun 2025 13:10
Operator : JC\MD
Sample : Q2414-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Jun 27 23:43:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 27 05:55:21 2025
Response via : Initial Calibration

