

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
 Data File : VN087217.D
 Acq On : 27 Jun 2025 13:31
 Operator : JC\MD
 Sample : Q2415-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Jun 27 23:44:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 27 05:55:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	164411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	341041	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	303735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	146320	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	134944	57.034	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.060%
35) Dibromofluoromethane	8.171	113	103447	46.918	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.840%
50) Toluene-d8	10.565	98	426341	46.752	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.500%
62) 4-Bromofluorobenzene	12.847	95	144440	45.647	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.300%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	45711	28.317	ug/l	96
18) Methyl Acetate	5.048	43	26136	8.323	ug/l	99
20) Methylene Chloride	5.295	84	9398	2.726	ug/l #	91
37) Ethyl Acetate	7.565	43	16951	1.865	ug/l	100
43) Isopropyl Acetate	8.700	43	11495	1.936	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
Data File : VN087217.D
Acq On : 27 Jun 2025 13:31
Operator : JC\MD
Sample : Q2415-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Jun 27 23:44:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 27 05:55:21 2025
Response via : Initial Calibration

