

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
 Data File : VN087219.D
 Acq On : 27 Jun 2025 14:12
 Operator : JC\MD
 Sample : Q2420-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11933

Quant Time: Jun 27 23:44:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 27 05:55:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	127267	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	301424	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	321898	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	147755	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	103060	56.271	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.540%
35) Dibromofluoromethane	8.171	113	74022	37.985	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	75.960%
50) Toluene-d8	10.565	98	380879	47.257	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.520%
62) 4-Bromofluorobenzene	12.847	95	146187	52.272	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.540%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	26335	17.080	ug/l	94
18) Methyl Acetate	5.041	43	21655	8.981	ug/l	100
20) Methylene Chloride	5.300	84	10967	5.131	ug/l	91
25) 2-Butanone	7.500	43	9161	6.597	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
Data File : VN087219.D
Acq On : 27 Jun 2025 14:12
Operator : JC\MD
Sample : Q2420-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11933

Quant Time: Jun 27 23:44:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 27 05:55:21 2025
Response via : Initial Calibration

