

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
 Data File : VN087220.D
 Acq On : 27 Jun 2025 14:32
 Operator : JC\MD
 Sample : Q2429-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Jun 27 23:45:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 27 05:55:21 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	187998	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	375149	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	349381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	140321	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	141233	52.203	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.400%
35) Dibromofluoromethane	8.171	113	111804	46.098	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.200%
50) Toluene-d8	10.565	98	443353	44.198	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.400%
62) 4-Bromofluorobenzene	12.847	95	146938	42.215	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	84.420%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	29182	8.908	ug/l	92
18) Methyl Acetate	5.047	43	23799	6.420	ug/l	97
20) Methylene Chloride	5.300	84	7846	1.447	ug/l	97
37) Ethyl Acetate	7.571	43	15775	1.207	ug/l	96
43) Isopropyl Acetate	8.694	43	10500	1.608	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062725\
Data File : VN087220.D
Acq On : 27 Jun 2025 14:32
Operator : JC\MD
Sample : Q2429-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Jun 27 23:45:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 27 05:55:21 2025
Response via : Initial Calibration

